

	文件类型	技术文件	文件版本	V11.3.0	
	项目代号	/	页码	第0页 共7页	
	文件编号	GL DM-56-03-en			
文件名称： <p style="text-align: center;">CONTROL DM-3D B0825 2025-11-10</p> 适用范围： 公司所有三分类仪器					
编制：刘金盈 日期：2025.07.18		审核：冯会娟 日期：2025.07.18		批准：程军 日期：2025.07.18	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.3.0	/	新建	刘金盈	程军	2025.07.18
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DH31 DH33 DH36 (Technical File Version A5.0 or higher) DH31 Vet DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	1.90 ± 0.60	8.10 ± 1.00	22.10 ± 2.50
	Lym%	%	61.4 ± 9.0	30.3 ± 8.0	13.7 ± 8.0
	Gran%	%	29.3 ± 9.0	62.8 ± 8.0	82.2 ± 8.0
	Mid%	%	9.3 ± 9.3	6.9 ± 6.0	4.1 ± 4.0
	Lym#	$\times 10^9/L$	1.16 ± 0.30	2.45 ± 0.70	3.02 ± 2.00
	Gran#	$\times 10^9/L$	0.56 ± 0.30	5.09 ± 0.70	18.17 ± 2.00
	Mid#	$\times 10^9/L$	0.18 ± 0.18	0.56 ± 0.50	0.91 ± 0.80
	RBC	$\times 10^{12}/L$	2.21 ± 0.24	4.65 ± 0.24	5.66 ± 0.50
	HGB	g/L	57 ± 6	136 ± 6	184 ± 8
	HCT	%	18.1 ± 2.0	42.7 ± 3.0	55.5 ± 4.0
	MCV	fL	82.1 ± 5.0	91.9 ± 5.0	98.0 ± 5.0
	MCH	pg	25.6 ± 2.5	29.2 ± 2.5	32.0 ± 2.5
	MCHC	g/L	315 ± 30	319 ± 30	330 ± 30
	RDW-CV	%	17.9 ± 3.0	17.6 ± 3.0	18.6 ± 3.0
	RDW-SD	fL	50.1 ± 8.0	55.4 ± 8.0	61.6 ± 8.0
	PLT	$\times 10^9/L$	69 ± 20	266 ± 40	539 ± 60
	MPV	fL	8.9 ± 3.0	8.5 ± 3.0	8.7 ± 3.0
	PDW	fL	11.0 ± 3.0	10.3 ± 3.0	10.8 ± 3.0
	PCT	%	0.062 ± 0.062	0.227 ± 0.200	0.476 ± 0.200
	P-LCR	%	20.1 ± 10.0	17.3 ± 10.0	19.7 ± 10.0
P-LCC	$\times 10^9/L$	14 ± 14	47 ± 25	108 ± 30	
PDW	/	15.9 ± 3.0	16.0 ± 3.0	15.9 ± 3.0	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0825L



B0825N





B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DH21 DH23 DH25 DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	1.90 ± 0.60	7.80 ± 1.00	21.40 ± 2.50
	Lym%	%	61.1 ± 9.0	29.6 ± 8.0	13.3 ± 8.0
	Gran%	%	29.7 ± 9.0	65.5 ± 8.0	81.0 ± 8.0
	Mid%	%	9.2 ± 9.2	4.9 ± 4.9	5.7 ± 5.0
	Lym#	$\times 10^9/L$	1.17 ± 0.30	2.31 ± 0.70	2.85 ± 2.00
	Gran#	$\times 10^9/L$	0.56 ± 0.30	5.11 ± 0.70	17.33 ± 2.00
	Mid#	$\times 10^9/L$	0.17 ± 0.17	0.38 ± 0.38	1.22 ± 1.10
	RBC	$\times 10^{12}/L$	2.21 ± 0.24	4.47 ± 0.24	5.44 ± 0.50
	HGB	g/L	56 ± 6	134 ± 6	183 ± 8
	HCT	%	17.2 ± 2.0	38.0 ± 3.0	49.3 ± 4.0
	MCV	fL	77.9 ± 5.0	85.0 ± 5.0	90.7 ± 5.0
	MCH	pg	25.1 ± 2.5	29.5 ± 2.5	33.0 ± 2.5
	MCHC	g/L	330 ± 30	352 ± 30	370 ± 30
	RDW-CV	%	17.1 ± 3.0	15.6 ± 3.0	15.7 ± 3.0
	RDW-SD	fL	48.9 ± 8.0	48.9 ± 8.0	51.4 ± 8.0
	PLT	$\times 10^9/L$	64 ± 20	259 ± 40	520 ± 60
	MPV	fL	10.0 ± 3.0	9.2 ± 3.0	9.3 ± 3.0
	PDW	fL	10.4 ± 3.0	10.0 ± 3.0	10.4 ± 3.0
	PCT	%	0.063 ± 0.063	0.236 ± 0.200	0.480 ± 0.200
P-LCR	%	26.8 ± 10.0	22.0 ± 10.0	23.3 ± 10.0	
P-LCC	$\times 10^9/L$	17 ± 15	56 ± 25	120 ± 30	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0825L



B0825N





B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DH20 DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	1.90 ± 0.60	8.10 ± 1.00	22.00 ± 2.50
	Lym%	%	60.7 ± 9.0	29.2 ± 8.0	13.6 ± 8.0
	Gran%	%	29.4 ± 9.0	64.9 ± 8.0	82.1 ± 8.0
	Mid%	%	9.9 ± 9.9	5.9 ± 5.9	4.3 ± 4.0
	Lym#	$\times 10^9/L$	1.15 ± 0.30	2.36 ± 0.70	2.99 ± 2.00
	Gran#	$\times 10^9/L$	0.56 ± 0.30	5.26 ± 0.70	18.06 ± 2.00
	Mid#	$\times 10^9/L$	0.19 ± 0.19	0.48 ± 0.48	0.95 ± 0.85
	RBC	$\times 10^{12}/L$	2.25 ± 0.24	4.72 ± 0.24	5.77 ± 0.50
	HGB	g/L	58 ± 6	137 ± 6	186 ± 8
	HCT	%	18.7 ± 2.0	42.1 ± 3.0	54.2 ± 4.0
	MCV	fL	83.3 ± 5.0	89.1 ± 5.0	93.9 ± 5.0
	MCH	pg	25.8 ± 2.5	28.1 ± 2.5	30.5 ± 2.5
	MCHC	g/L	310 ± 30	325 ± 30	331 ± 30
	RDW-CV	%	15.6 ± 3.0	14.3 ± 3.0	14.7 ± 3.0
	RDW-SD	fL	53.0 ± 8.0	53.5 ± 8.0	57.6 ± 8.0
	PLT	$\times 10^9/L$	66 ± 20	266 ± 40	543 ± 60
	MPV	fL	8.6 ± 3.0	8.4 ± 3.0	8.7 ± 3.0
	PDW	fL	13.3 ± 3.0	12.8 ± 3.0	13.6 ± 3.0
	PCT	%	0.058 ± 0.058	0.227 ± 0.200	0.483 ± 0.200
P-LCR	%	31.6 ± 10.0	29.8 ± 10.0	32.6 ± 10.0	
P-LCC	$\times 10^9/L$	21 ± 15	81 ± 25	181 ± 30	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0825L



B0825N





B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DH31 Vet DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	1.80 ± 0.60	7.40 ± 1.00	20.10 ± 2.50
	Lym%	%	58.3 ± 9.0	28.5 ± 8.0	12.9 ± 8.0
	Gran%	%	30.2 ± 9.0	65.3 ± 8.0	84.5 ± 8.0
	Mid%	%	11.5 ± 10.0	6.2 ± 6.0	2.6 ± 2.6
	Lym#	$\times 10^9/L$	1.05 ± 0.30	2.11 ± 0.70	2.60 ± 2.00
	Gran#	$\times 10^9/L$	0.54 ± 0.30	4.83 ± 0.70	16.98 ± 2.00
	Mid#	$\times 10^9/L$	0.21 ± 0.20	0.46 ± 0.46	0.52 ± 0.50
	RBC	$\times 10^{12}/L$	2.15 ± 0.30	4.35 ± 0.40	5.37 ± 0.50
	HGB	g/L	53 ± 6	125 ± 6	168 ± 8
	HCT	%	18.0 ± 2.0	39.8 ± 3.0	52.0 ± 4.0
	MCV	fL	83.7 ± 5.0	91.4 ± 5.0	96.9 ± 5.0
	MCH	pg	24.7 ± 2.5	29.6 ± 2.5	31.3 ± 2.5
	MCHC	g/L	310 ± 30	325 ± 30	323 ± 30
	RDW-CV	%	16.3 ± 3.0	14.9 ± 3.0	15.2 ± 3.0
	RDW-SD	fL	53.9 ± 8.0	53.8 ± 8.0	57.4 ± 8.0
	PLT	$\times 10^9/L$	63 ± 20	244 ± 40	517 ± 60
	MPV	fL	9.7 ± 3.0	8.9 ± 3.0	9.2 ± 3.0
	PDW	fL	10.4 ± 3.0	9.7 ± 3.0	10.2 ± 3.0
PCT	%	0.064 ± 0.064	0.247 ± 0.200	0.437 ± 0.200	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0825L




B0825N




B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	1.80 ± 0.60	7.40 ± 1.00	20.40 ± 2.50
	Lym%	%	66.3 ± 9.0	33.7 ± 8.0	15.6 ± 8.0
	Gran%	%	30.6 ± 9.0	61.8 ± 8.0	79.7 ± 8.0
	Mid%	%	3.1 ± 3.1	4.5 ± 4.5	4.7 ± 4.0
	Lym#	$\times 10^9/L$	1.19 ± 0.30	2.50 ± 0.70	3.18 ± 2.00
	Gran#	$\times 10^9/L$	0.55 ± 0.30	4.57 ± 0.70	16.26 ± 2.00
	Mid#	$\times 10^9/L$	0.06 ± 0.06	0.33 ± 0.33	0.96 ± 0.80
	RBC	$\times 10^{12}/L$	2.25 ± 0.24	4.55 ± 0.24	5.56 ± 0.50
	HGB	g/L	56 ± 6	132 ± 6	179 ± 8
	HCT	%	18.2 ± 2.0	40.2 ± 3.0	52.5 ± 4.0
	MCV	fL	80.7 ± 5.0	88.3 ± 5.0	94.4 ± 5.0
	MCH	pg	24.7 ± 2.5	29.1 ± 2.5	32.2 ± 2.5
	MCHC	g/L	311 ± 30	330 ± 30	345 ± 30
	RDW-CV	%	17.3 ± 3.0	15.8 ± 3.0	16.1 ± 3.0
	RDW-SD	fL	52.0 ± 8.0	51.8 ± 8.0	55.2 ± 8.0
	PLT	$\times 10^9/L$	65 ± 20	253 ± 40	524 ± 60
	MPV	fL	9.4 ± 3.0	9.0 ± 3.0	9.2 ± 3.0
	PDW	fL	12.5 ± 3.0	11.7 ± 3.0	12.2 ± 3.0
	PCT	%	0.062 ± 0.062	0.231 ± 0.200	0.488 ± 0.200
	P-LCR	%	28.5 ± 10.0	25.0 ± 10.0	26.9 ± 10.0
P-LCC	$\times 10^9/L$	19 ± 15	64 ± 25	142 ± 30	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0825L



B0825N





B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DP-H10 DP-H12 DP-H13 DP-H15 DP-H16 DP-H17 (Technical File Version A3.3 or higher)	WBC	$\times 10^9/L$	2.20 ± 0.60	8.40 ± 1.00	22.20 ± 2.50
	Lym%	%	64.0 ± 9.0	31.4 ± 8.0	14.3 ± 8.0
	Gran%	%	31.2 ± 9.0	65.3 ± 8.0	82.4 ± 8.0
	Mid%	%	4.8 ± 4.8	3.3 ± 3.3	3.3 ± 3.3
	Lym#	$\times 10^9/L$	1.40 ± 0.30	2.63 ± 0.70	3.18 ± 2.00
	Gran#	$\times 10^9/L$	0.69 ± 0.30	5.49 ± 0.70	18.29 ± 2.00
	Mid#	$\times 10^9/L$	0.11 ± 0.11	0.28 ± 0.28	0.73 ± 0.70
	RBC	$\times 10^{12}/L$	2.45 ± 0.24	4.94 ± 0.24	6.09 ± 0.50
	HGB	g/L	59 ± 6	136 ± 6	183 ± 8
	HCT	%	19.4 ± 2.0	42.2 ± 3.0	54.6 ± 4.0
	MCV	fL	79.2 ± 5.0	85.5 ± 5.0	89.6 ± 5.0
	MCH	pg	24.0 ± 2.5	27.6 ± 2.5	30.1 ± 2.5
	MCHC	g/L	308 ± 30	327 ± 30	341 ± 30
	RDW-CV	%	16.1 ± 3.0	15.3 ± 3.0	15.5 ± 3.0
	RDW-SD	fL	49.2 ± 8.0	50.9 ± 8.0	53.5 ± 8.0
	PLT	$\times 10^9/L$	48 ± 20	240 ± 40	485 ± 60
	MPV	fL	9.4 ± 3.0	9.0 ± 3.0	9.3 ± 3.0
	PDW	fL	12.2 ± 3.0	11.1 ± 3.0	11.9 ± 3.0
	PCT	%	0.044 ± 0.044	0.215 ± 0.200	0.452 ± 0.200
	P-LCR	%	23.8 ± 10.0	22.0 ± 10.0	24.2 ± 10.0
P-LCC	$\times 10^9/L$	12 ± 12	53 ± 25	118 ± 30	

【NOTE】

- The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the controls back into the refrigerator to prevent contamination and evaporation.
- P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



B0825L



B0825N





B0825H

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-07-14

 2025-11-10

Applicable Instruments	Parameter	Unit	LOT B0825L	LOT B0825N	LOT B0825H
DYMIND DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	1.80 ± 0.80	7.40 ± 2.00	20.10 ± 4.50
	Lym%	%	59.3 ± 10.0	28.5 ± 8.0	12.9 ± 8.0
	Gran%	%	30.2 ± 12.0	65.3 ± 8.0	84.5 ± 8.0
	Mid%	%	10.5 ± 10.0	6.2 ± 5.0	2.6 ± 2.6
	Lym#	$\times 10^9/L$	1.07 ± 0.60	2.11 ± 1.00	2.60 ± 2.00
	Gran#	$\times 10^9/L$	0.54 ± 0.50	4.83 ± 1.80	16.98 ± 5.00
	Mid#	$\times 10^9/L$	0.19 ± 0.19	0.46 ± 0.46	0.52 ± 0.50
	RBC	$\times 10^{12}/L$	2.16 ± 0.50	4.37 ± 1.20	5.40 ± 1.20
	HGB	g/L	63 ± 15	135 ± 30	180 ± 40
	HCT	%	18.0 ± 5.0	39.9 ± 10.0	52.0 ± 15.0
	MCV	fL	83.3 ± 12.0	91.2 ± 8.0	96.3 ± 8.0
	MCH	pg	29.5 ± 3.0	31.3 ± 3.0	34.0 ± 3.0
	MCHC	g/L	359 ± 45	353 ± 45	362 ± 45
	RDW-CV	%	15.7 ± 3.0	15.2 ± 3.0	14.9 ± 3.0
	RDW-SD	fL	36.8 ± 8.0	38.5 ± 8.0	39.4 ± 8.0
	PLT	$\times 10^9/L$	74 ± 40	249 ± 80	525 ± 120
	MPV	fL	10.7 ± 4.0	9.9 ± 3.0	10.2 ± 3.0
	PDW	fL	10.4 ± 9.0	7.8 ± 3.0	7.9 ± 3.0
PCT	%	0.069 ± 0.069	0.249 ± 0.249	0.537 ± 0.400	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



B0825L



B0825N



B0825H