

	文件类型	技术文件	文件版本	V11.9.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-05-en			
文件名称： <p style="text-align: center;">CALIBRATOR N0925 2025-10-10</p> 适用范围： 公司所有三分类仪器					
编制：刘金盈 日期：2025.08.21		审核：冯会娟 日期：2025.08.21		批准：程军 日期：2025.08.21	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.9.0	/	新建	刘金盈	程军	2025.08.21
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

DM-CAL

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0925
-----	-------

📅	2025-08-12
---	------------

🕒	2025-10-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.40
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	129
	MCV	fL	90.6
	PLT	$\times 10^9/L$	247
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.33
	RBC	$\times 10^{12}/L$	4.33
	HGB	g/L	127
	MCV	fL	83.7
	PLT	$\times 10^9/L$	244
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.68
	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	131
	MCV	fL	88.4
	PLT	$\times 10^9/L$	247
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.01
	RBC	$\times 10^{12}/L$	4.28
	HGB	g/L	122
	MCV	fL	90.7
	PLT	$\times 10^9/L$	233

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0925
-----	-------

📅	2025-08-12
---	------------

🕒	2025-10-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.03
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	126
	MCV	fL	86.6
	PLT	$\times 10^9/L$	240
DP-H10, DP-H12 DP-H13, DP-H15 DP-H16, DP-H17 (Technical File Version A3.3 or higher)	WBC	$\times 10^9/L$	9.95
	RBC	$\times 10^{12}/L$	4.70
	HGB	g/L	137
	MCV	fL	88.5
	PLT	$\times 10^9/L$	263
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.01
	RBC	$\times 10^{12}/L$	4.28
	HGB	g/L	132
	MCV	fL	90.7
	PLT	$\times 10^9/L$	233

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.