



	文件类型	技术文件	文件版本	V5.9.1	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-17-en			
文件名称： <p style="text-align: center;"><b>CONTROL DM-RET RET0925 2025-11-25</b></p> 适用范围： 1116项目、7003项目					
编制：刘金盈 日期：2025.09.22		审核：冯会娟 日期：2025.09.22		批准：程军 日期：2025.09.22	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V5.9.0	/	新建	刘金盈	程军	2025.09.05
V5.9.1	/	更新1116二维码	刘金盈	程军	2025.09.22
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-08-28

 2025-11-25

Applicable Instruments	Parameter	Unit	LOT RET0925L	LOT RET0925N	LOT RET0925H
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0401 $\pm 0.0201$	0.1929 $\pm 0.1157$	0.2135 $\pm 0.1068$
	RET%	%	0.74 $\pm 0.37$	4.31 $\pm 1.72$	9.34 $\pm 3.74$
	LFR	%	89.6 $\pm 27.0$	89.8 $\pm 27.0$	89.1 $\pm 27.0$
	MFR	%	8.3 $\pm 12.0$	8.1 $\pm 12.0$	8.9 $\pm 13.0$
	HFR	%	2.1 $\pm 3.0$	2.1 $\pm 3.0$	2.0 $\pm 3.0$
	IRF	%	10.4 $\pm 16.0$	10.2 $\pm 15.0$	10.9 $\pm 16.0$
	RHE	pg	16.8 $\pm 7.0$	15.0 $\pm 6.0$	16.2 $\pm 6.0$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0925L




RET0925N




RET0925H

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-08-28

 2025-11-25

Applicable Instruments	Parameter	Unit	LOT RET0925L	LOT RET0925N	LOT RET0925H
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0149 $\pm 0.0119$	0.0899 $\pm 0.0450$	0.1025 $\pm 0.0513$
	RET%	%	0.29 $\pm 0.23$	2.10 $\pm 1.05$	4.72 $\pm 2.36$

**【NOTE】**

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0925L



RET0925N



RET0925H