



	文件类型	技术文件	文件版本	V11.4.0	
	项目代号	/	页码	第0页 共7页	
	文件编号	GL DM-56-03-en			
文件名称： <p style="text-align: center;"><b>CONTROL DM-3D B1125 2026-02-10</b></p> 适用范围： 公司所有三分类仪器					
编制：刘金盈 日期：2025.10.22		审核：冯会娟 日期：2025.10.22		批准：程军 日期：2025.10.22	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.4.0	/	新建	刘金盈	程军	2025.10.22
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

# DM-3D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b> DH31 DH33 DH36 (Technical File Version A5.0 or higher) DH31 Vet DH36 Vet (Technical File Version A1.0 to A4.0)	<b>WBC</b>	$\times 10^9/L$	<b>2.00</b> $\pm 0.60$	<b>7.72</b> $\pm 1.00$	<b>21.68</b> $\pm 2.50$
	Lym%	%	59.4 $\pm 9.0$	32.1 $\pm 8.0$	12.0 $\pm 8.0$
	Gran%	%	30.3 $\pm 9.0$	61.7 $\pm 8.0$	81.5 $\pm 8.0$
	Mid%	%	10.3 $\pm 10.0$	6.2 $\pm 6.0$	6.5 $\pm 6.0$
	Lym#	$\times 10^9/L$	1.18 $\pm 0.30$	2.48 $\pm 0.70$	2.60 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.61 $\pm 0.30$	4.76 $\pm 0.70$	17.67 $\pm 2.00$
	Mid#	$\times 10^9/L$	0.21 $\pm 0.20$	0.48 $\pm 0.48$	1.41 $\pm 1.30$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.31</b> $\pm 0.24$	<b>4.63</b> $\pm 0.24$	<b>5.75</b> $\pm 0.50$
	<b>HGB</b>	g/L	<b>58</b> $\pm 6$	<b>135</b> $\pm 6$	<b>185</b> $\pm 8$
	HCT	%	18.4 $\pm 2.0$	42.7 $\pm 3.0$	56.9 $\pm 4.0$
	<b>MCV</b>	fL	<b>81.8</b> $\pm 5.0$	<b>92.2</b> $\pm 5.0$	<b>101.0</b> $\pm 5.0$
	MCH	pg	25.8 $\pm 2.5$	29.9 $\pm 2.5$	33.4 $\pm 2.5$
	MCHC	g/L	321 $\pm 30$	326 $\pm 30$	336 $\pm 30$
	RDW-CV	%	17.8 $\pm 3.0$	18.2 $\pm 3.0$	17.4 $\pm 3.0$
	RDW-SD	fL	49.0 $\pm 8.0$	54.7 $\pm 8.0$	58.6 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>67</b> $\pm 20$	<b>249</b> $\pm 40$	<b>507</b> $\pm 60$
	MPV	fL	7.6 $\pm 3.0$	8.6 $\pm 3.0$	8.5 $\pm 3.0$
	PDW	fL	8.8 $\pm 3.0$	10.5 $\pm 3.0$	10.4 $\pm 3.0$
	PCT	%	0.052 $\pm 0.052$	0.219 $\pm 0.200$	0.439 $\pm 0.200$
	P-LCR	%	12.0 $\pm 10.0$	18.2 $\pm 10.0$	17.9 $\pm 10.0$
P-LCC	$\times 10^9/L$	12 $\pm 12$	46 $\pm 25$	93 $\pm 30$	
PDW	/	15.8 $\pm 3.0$	16.2 $\pm 3.0$	15.9 $\pm 3.0$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1125L




B1125N




B1125H

# DM-3D

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b>  DH21 DH23 DH25 DH26 (Technical File Version A1.0 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>2.02</b> $\pm 0.60$	<b>7.78</b> $\pm 1.00$	<b>21.64</b> $\pm 2.50$
	Lym%	%	58.3 $\pm 9.0$	31.1 $\pm 8.0$	11.5 $\pm 8.0$
	Gran%	%	31.7 $\pm 9.0$	64.4 $\pm 8.0$	83.6 $\pm 8.0$
	Mid%	%	10.0 $\pm 10.0$	4.5 $\pm 4.5$	4.9 $\pm 4.9$
	Lym#	$\times 10^9/L$	1.18 $\pm 0.30$	2.42 $\pm 0.70$	2.49 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.64 $\pm 0.30$	5.01 $\pm 0.70$	18.09 $\pm 2.00$
	Mid#	$\times 10^9/L$	0.20 $\pm 0.20$	0.35 $\pm 0.35$	1.06 $\pm 1.06$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.29</b> $\pm 0.24$	<b>4.48</b> $\pm 0.24$	<b>5.48</b> $\pm 0.50$
	<b>HGB</b>	g/L	<b>58</b> $\pm 6$	<b>137</b> $\pm 6$	<b>188</b> $\pm 8$
	HCT	%	17.7 $\pm 2.0$	38.4 $\pm 3.0$	50.7 $\pm 4.0$
	<b>MCV</b>	fL	<b>77.2</b> $\pm 5.0$	<b>85.8</b> $\pm 5.0$	<b>92.6</b> $\pm 5.0$
	MCH	pg	25.1 $\pm 2.5$	30.4 $\pm 2.5$	34.2 $\pm 2.5$
	MCHC	g/L	338 $\pm 30$	362 $\pm 30$	380 $\pm 30$
	RDW-CV	%	17.2 $\pm 3.0$	16.4 $\pm 3.0$	15.4 $\pm 3.0$
	RDW-SD	fL	48.0 $\pm 8.0$	50.6 $\pm 8.0$	51.1 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>62</b> $\pm 20$	<b>248</b> $\pm 40$	<b>506</b> $\pm 60$
	MPV	fL	8.6 $\pm 3.0$	9.2 $\pm 3.0$	9.2 $\pm 3.0$
	PDW	fL	8.8 $\pm 3.0$	10.1 $\pm 3.0$	10.1 $\pm 3.0$
	PCT	%	0.053 $\pm 0.053$	0.227 $\pm 0.200$	0.466 $\pm 0.200$
P-LCR	%	17.7 $\pm 10.0$	21.8 $\pm 10.0$	22.4 $\pm 10.0$	
P-LCC	$\times 10^9/L$	11 $\pm 11$	54 $\pm 25$	113 $\pm 30$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1125L



B1125N





B1125H

# DM-3D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b>  DH20 DH22 (Technical File Version B1.0 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>2.00</b> $\pm 0.60$	<b>7.85</b> $\pm 1.00$	<b>21.64</b> $\pm 2.50$
	Lym%	%	57.4 $\pm 9.0$	31.5 $\pm 8.0$	12.0 $\pm 8.0$
	Gran%	%	32.0 $\pm 9.0$	63.6 $\pm 8.0$	82.9 $\pm 8.0$
	Mid%	%	10.6 $\pm 10.0$	4.9 $\pm 4.9$	5.1 $\pm 4.0$
	Lym#	$\times 10^9/L$	1.15 $\pm 0.30$	2.48 $\pm 0.70$	2.60 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.64 $\pm 0.30$	4.99 $\pm 0.70$	17.94 $\pm 2.00$
	Mid#	$\times 10^9/L$	0.21 $\pm 0.20$	0.38 $\pm 0.38$	1.10 $\pm 0.50$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.36</b> $\pm 0.24$	<b>4.75</b> $\pm 0.24$	<b>5.83</b> $\pm 0.50$
	<b>HGB</b>	g/L	<b>59</b> $\pm 6$	<b>138</b> $\pm 6$	<b>189</b> $\pm 8$
	HCT	%	19.6 $\pm 2.0$	42.9 $\pm 3.0$	56.3 $\pm 4.0$
	<b>MCV</b>	fL	<b>83.1</b> $\pm 5.0$	<b>90.3</b> $\pm 5.0$	<b>96.6</b> $\pm 5.0$
	MCH	pg	26.2 $\pm 2.5$	30.3 $\pm 2.5$	33.4 $\pm 2.5$
	MCHC	g/L	324 $\pm 30$	343 $\pm 30$	351 $\pm 30$
	RDW-CV	%	15.9 $\pm 3.0$	15.3 $\pm 3.0$	14.1 $\pm 3.0$
	RDW-SD	fL	53.5 $\pm 8.0$	57.0 $\pm 8.0$	56.7 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>63</b> $\pm 20$	<b>248</b> $\pm 40$	<b>503</b> $\pm 60$
	MPV	fL	7.5 $\pm 3.0$	8.5 $\pm 3.0$	8.5 $\pm 3.0$
	PDW	fL	10.2 $\pm 3.0$	13.2 $\pm 3.0$	13.2 $\pm 3.0$
	PCT	%	0.049 $\pm 0.049$	0.216 $\pm 0.200$	0.438 $\pm 0.200$
P-LCR	%	21.3 $\pm 10.0$	31.1 $\pm 10.0$	30.8 $\pm 10.0$	
P-LCC	$\times 10^9/L$	14 $\pm 14$	79 $\pm 25$	164 $\pm 30$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1125L




B1125N




B1125H

# DM-3D

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b>  DH31 Vet DH36 Vet (Technical File Version A5.0 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>1.81</b> $\pm 0.60$	<b>7.16</b> $\pm 1.00$	<b>19.86</b> $\pm 2.50$
	Lym%	%	55.7 $\pm 9.0$	29.9 $\pm 8.0$	11.1 $\pm 8.0$
	Gran%	%	31.2 $\pm 9.0$	63.6 $\pm 8.0$	84.8 $\pm 8.0$
	Mid%	%	13.1 $\pm 10.0$	6.5 $\pm 6.0$	4.1 $\pm 4.0$
	Lym#	$\times 10^9/L$	1.01 $\pm 0.30$	2.14 $\pm 0.70$	2.21 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.56 $\pm 0.30$	4.55 $\pm 0.70$	16.84 $\pm 2.00$
	Mid#	$\times 10^9/L$	0.24 $\pm 0.20$	0.47 $\pm 0.47$	0.81 $\pm 0.80$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.23</b> $\pm 0.30$	<b>4.38</b> $\pm 0.40$	<b>5.39</b> $\pm 0.50$
	<b>HGB</b>	g/L	<b>54</b> $\pm 6$	<b>126</b> $\pm 6$	<b>172</b> $\pm 8$
	HCT	%	18.5 $\pm 2.0$	40.1 $\pm 3.0$	53.6 $\pm 4.0$
	<b>MCV</b>	fL	<b>82.9</b> $\pm 5.0$	<b>91.6</b> $\pm 5.0$	<b>99.4</b> $\pm 5.0$
	MCH	pg	25.4 $\pm 2.5$	30.0 $\pm 2.5$	33.4 $\pm 2.5$
	MCHC	g/L	317 $\pm 30$	330 $\pm 30$	340 $\pm 30$
	RDW-CV	%	16.3 $\pm 3.0$	15.5 $\pm 3.0$	14.3 $\pm 3.0$
	RDW-SD	fL	53.6 $\pm 8.0$	56.1 $\pm 8.0$	56.1 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>65</b> $\pm 20$	<b>241</b> $\pm 40$	<b>505</b> $\pm 60$
	MPV	fL	8.7 $\pm 3.0$	9.3 $\pm 3.0$	9.2 $\pm 3.0$
	PDW	fL	8.6 $\pm 3.0$	9.7 $\pm 3.0$	9.7 $\pm 3.0$
PCT	%	0.060 $\pm 0.060$	0.243 $\pm 0.200$	0.464 $\pm 0.200$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1125L



B1125N





B1125H

# DM-3D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b>  DH36X (Technical File Version A1.0 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>1.90</b> $\pm 0.60$	<b>7.36</b> $\pm 1.00$	<b>20.20</b> $\pm 2.50$
	Lym%	%	64.9 $\pm 9.0$	34.9 $\pm 8.0$	14.6 $\pm 8.0$
	Gran%	%	31.9 $\pm 9.0$	62.9 $\pm 8.0$	81.7 $\pm 8.0$
	Mid%	%	3.2 $\pm 3.2$	2.2 $\pm 2.2$	3.7 $\pm 3.7$
	Lym#	$\times 10^9/L$	1.23 $\pm 0.30$	2.57 $\pm 0.70$	2.95 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.61 $\pm 0.30$	4.63 $\pm 0.70$	16.50 $\pm 2.00$
	Mid#	$\times 10^9/L$	0.06 $\pm 0.06$	0.16 $\pm 0.16$	0.75 $\pm 0.75$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.32</b> $\pm 0.24$	<b>4.54</b> $\pm 0.24$	<b>5.54</b> $\pm 0.50$
	<b>HGB</b>	g/L	<b>57</b> $\pm 6$	<b>132</b> $\pm 6$	<b>180</b> $\pm 8$
	HCT	%	18.6 $\pm 2.0$	40.3 $\pm 3.0$	53.2 $\pm 4.0$
	<b>MCV</b>	fL	<b>80.1</b> $\pm 5.0$	<b>88.8</b> $\pm 5.0$	<b>96.0</b> $\pm 5.0$
	MCH	pg	25.1 $\pm 2.5$	29.7 $\pm 2.5$	33.2 $\pm 2.5$
	MCHC	g/L	321 $\pm 30$	339 $\pm 30$	352 $\pm 30$
	RDW-CV	%	17.4 $\pm 3.0$	16.5 $\pm 3.0$	15.3 $\pm 3.0$
	RDW-SD	fL	51.1 $\pm 8.0$	53.5 $\pm 8.0$	53.3 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>68</b> $\pm 20$	<b>249</b> $\pm 40$	<b>504</b> $\pm 60$
	MPV	fL	8.3 $\pm 3.0$	9.1 $\pm 3.0$	9.0 $\pm 3.0$
	PDW	fL	10.2 $\pm 3.0$	11.9 $\pm 3.0$	11.7 $\pm 3.0$
	PCT	%	0.057 $\pm 0.057$	0.230 $\pm 0.200$	0.461 $\pm 0.200$
	P-LCR	%	19.9 $\pm 10.0$	25.6 $\pm 10.0$	25.1 $\pm 10.0$
P-LCC	$\times 10^9/L$	14 $\pm 14$	65 $\pm 25$	129 $\pm 30$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1125L



B1125N





B1125H

# DM-3D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b> DP-H10 DP-H12 DP-H13 DP-H15 DP-H16 DP-H17 (Technical File Version A3.3 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>2.10</b> $\pm 0.80$	<b>8.00</b> $\pm 1.20$	<b>21.80</b> $\pm 3.00$
	Lym%	%	62.0 $\pm 10.0$	33.5 $\pm 8.0$	12.8 $\pm 8.0$
	Gran%	%	32.6 $\pm 12.0$	63.9 $\pm 8.0$	83.9 $\pm 8.0$
	Mid%	%	5.4 $\pm 5.4$	2.6 $\pm 2.6$	3.3 $\pm 3.3$
	Lym#	$\times 10^9/L$	1.31 $\pm 0.60$	2.68 $\pm 1.00$	2.79 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.68 $\pm 0.50$	5.11 $\pm 1.80$	18.29 $\pm 5.00$
	Mid#	$\times 10^9/L$	0.11 $\pm 0.11$	0.21 $\pm 0.21$	0.72 $\pm 0.72$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.51</b> $\pm 0.30$	<b>4.88</b> $\pm 0.60$	<b>6.01</b> $\pm 0.60$
	<b>HGB</b>	g/L	<b>62</b> $\pm 8$	<b>141</b> $\pm 10$	<b>194</b> $\pm 10$
	HCT	%	19.8 $\pm 3.0$	41.5 $\pm 5.0$	55.4 $\pm 7.0$
	<b>MCV</b>	fL	<b>78.8</b> $\pm 12.0$	<b>85.1</b> $\pm 8.0$	<b>92.2</b> $\pm 8.0$
	MCH	pg	24.8 $\pm 6.0$	28.9 $\pm 3.5$	32.3 $\pm 3.5$
	MCHC	g/L	321 $\pm 70$	344 $\pm 70$	356 $\pm 70$
	RDW-CV	%	16.8 $\pm 3.0$	16.1 $\pm 3.0$	14.8 $\pm 3.0$
	RDW-SD	fL	50.9 $\pm 8.0$	53.1 $\pm 8.0$	53.7 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>52</b> $\pm 40$	<b>264</b> $\pm 60$	<b>519</b> $\pm 80$
	MPV	fL	8.2 $\pm 4.0$	8.8 $\pm 3.0$	9.1 $\pm 3.0$
	PDW	fL	10.4 $\pm 9.0$	10.7 $\pm 3.0$	11.3 $\pm 3.0$
	PCT	%	0.043 $\pm 0.043$	0.232 $\pm 0.232$	0.472 $\pm 0.400$
	P-LCR	%	17.6 $\pm 15.0$	21.6 $\pm 15.0$	23.0 $\pm 15.0$
P-LCC	$\times 10^9/L$	9 $\pm 9$	57 $\pm 30$	120 $\pm 40$	

**【NOTE】**

- The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the controls back into the refrigerator to prevent contamination and evaporation.
- P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



B1125L




B1125N




B1125H

# DM-3D

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-10-14

 2026-02-10

Applicable Instruments	Parameter	Unit	LOT B1125L	LOT B1125N	LOT B1125H
<b>DYMIND</b>  DP-H10 Vet (Technical File Version A1.0 or higher)	<b>WBC</b>	$\times 10^9/L$	<b>1.81</b> $\pm 0.80$	<b>7.16</b> $\pm 2.00$	<b>19.86</b> $\pm 4.50$
	Lym%	%	55.7 $\pm 10.0$	29.9 $\pm 8.0$	11.1 $\pm 8.0$
	Gran%	%	31.2 $\pm 12.0$	63.6 $\pm 8.0$	84.8 $\pm 8.0$
	Mid%	%	13.1 $\pm 10.0$	6.5 $\pm 5.0$	4.1 $\pm 4.0$
	Lym#	$\times 10^9/L$	1.01 $\pm 0.60$	2.14 $\pm 1.00$	2.21 $\pm 2.00$
	Gran#	$\times 10^9/L$	0.56 $\pm 0.50$	4.55 $\pm 1.80$	16.84 $\pm 5.00$
	Mid#	$\times 10^9/L$	0.24 $\pm 0.24$	0.47 $\pm 0.47$	0.81 $\pm 0.80$
	<b>RBC</b>	$\times 10^{12}/L$	<b>2.23</b> $\pm 0.50$	<b>4.38</b> $\pm 1.20$	<b>5.39</b> $\pm 1.20$
	<b>HGB</b>	g/L	<b>64</b> $\pm 15$	<b>136</b> $\pm 30$	<b>182</b> $\pm 40$
	HCT	%	18.5 $\pm 5.0$	40.1 $\pm 10.0$	53.6 $\pm 15.0$
	<b>MCV</b>	fL	<b>84.1</b> $\pm 12.0$	<b>93.0</b> $\pm 8.0$	<b>99.4</b> $\pm 8.0$
	MCH	pg	30.0 $\pm 3.0$	32.0 $\pm 3.0$	35.1 $\pm 3.0$
	MCHC	g/L	370 $\pm 45$	353 $\pm 45$	360 $\pm 45$
	RDW-CV	%	15.5 $\pm 3.0$	14.8 $\pm 3.0$	13.5 $\pm 3.0$
	RDW-SD	fL	35.7 $\pm 8.0$	38.1 $\pm 8.0$	36.8 $\pm 8.0$
	<b>PLT</b>	$\times 10^9/L$	<b>65</b> $\pm 40$	<b>241</b> $\pm 80$	<b>495</b> $\pm 120$
	MPV	fL	8.7 $\pm 4.0$	9.3 $\pm 3.0$	9.2 $\pm 3.0$
	PDW	fL	8.6 $\pm 8.6$	9.7 $\pm 3.0$	9.7 $\pm 3.0$
PCT	%	0.060 $\pm 0.060$	0.243 $\pm 0.243$	0.464 $\pm 0.400$	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



B1125L



B1125N



B1125H