
	文件类型	技术文件	文件版本	V6.4.0	
	项目代号	/	页码	第0页 共1页	
	文件编号	GL DM-56-07-en			
文件名称： <p style="text-align: center;">Table of C-Reactive Protein Calibrator (CRP) Reference Value 2025091801</p> 适用范围： DM3113项目、DM3513项目					
编制：刘金盈 日期：2025.10.27		审核：冯会娟 日期：2025.10.27		批准：程军 日期：2025.10.27	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V6.4.0	/	新建	刘金盈	程军	2025.10.27
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

Table of C-Reactive Protein Calibrator (CRP) Reference Value

C-Reactive Protein Calibrator

LOT 2025091801

 2026-09-17

Applicable Instruments	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	DH71CRP, DH73CRP D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	DH-2000T series , DH-100T, DH-102T DH-103T, DH-105T DH-106T, DH-108T (Technical file version A2.0 or higher)	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)
Level (multi-point)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)
Calibrator 1	4.9	4.9	4.5	4.6	4.53	3.6	3.9
Calibrator 2	17.4	17.2	18.9	17.9	19.02	16.8	16.2
Calibrator 3	40.9	43.0	46.9	44.4	44.78	42.6	45.0
Calibrator 4	89.5	89.3	86.5	95.4	86.17	80.6	100.0
Calibrator 5	131.5	148.2	139.4	124.7	112.20	104.2	160.0
Calibrator 6	158.2	190.1	162.8	157.8	152.00	140.0	180.0
Applicable Instruments	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	DM79X, DM78X, DM72X (Technical file version A1.0 or higher)	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher)	DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher)	/	/	/
Level (multi-point)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	Reference Value (mg/L)	/	/	/
Calibrator 1	5.1	5.59	4.3	4.36	/	/	/
Calibrator 2	19.5	19.89	16.4	19.14	/	/	/
Calibrator 3	45.3	49.65	46.1	46.41	/	/	/
Calibrator 4	101.6	88.16	92.7	87.04	/	/	/
Calibrator 5	143.1	121.50	129.1	109.95	/	/	/
Calibrator 6	178.6	157.00	158.4	153.00	/	/	/