
	文件类型	技术文件	文件版本	V6.4.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-06-en			
文件名称： <p style="text-align: center;"><b>Table of C-Reactive Protein Control (CRP) Reference Value 2025091801</b></p> 适用范围： DM3113项目、DM3513项目					
编制：刘金盈 日期：2025.10.27		审核：冯会娟 日期：2025.10.27		批准：程军 日期：2025.10.27	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V6.4.0	/	新建	刘金盈	程军	2025.10.27
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2025091801


 2026-09-17

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	Control 1	<b>3.3</b>	±1.0	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	Control 1	<b>3.2</b>	±1.0
	Control 2	<b>14.1</b>	±4.2		Control 2	<b>14.3</b>	±4.3
	Control 3	<b>33.7</b>	±10.1		Control 3	<b>33.6</b>	±10.1
DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>2.7</b>	±0.8	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	Control 1	<b>2.8</b>	±0.8
	Control 2	<b>14.9</b>	±4.5		Control 2	<b>13.7</b>	±4.1
	Control 3	<b>39.4</b>	±11.8		Control 3	<b>35.7</b>	±10.7
DH-2000T series , DH-100T, DH-102T, DH-103T, DH-105T, DH-106T, DH-108T (Technical file version A2.0 or higher)	Control 1	<b>2.84</b>	±0.85	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	Control 1	<b>2.5</b>	±0.8
	Control 2	<b>14.53</b>	±4.36		Control 2	<b>12.9</b>	±3.9
	Control 3	<b>36.61</b>	±10.98		Control 3	<b>34.4</b>	±10.3
DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>2.5</b>	±0.8	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	Control 1	<b>3.3</b>	±1.0
	Control 2	<b>11.8</b>	±3.5		Control 2	<b>15.5</b>	±4.6
	Control 3	<b>33.4</b>	±10.0		Control 3	<b>38.2</b>	±11.5

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2025091801

 2026-09-17

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DM79X, DM78X, DM72X (Technical file version A1.0 or higher )	Control 1	<b>3.63</b>	±1.1	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher )	Control 1	<b>2.7</b>	±0.8
	Control 2	<b>15.97</b>	±4.8		Control 2	<b>12.5</b>	±3.8
	Control 3	<b>40.04</b>	±12.0		Control 3	<b>34.9</b>	±10.5
DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher )	Control 1	<b>3.06</b>	±0.92	/	/	/	/
	Control 2	<b>15.69</b>	±4.71	/	/	/	/
	Control 3	<b>38.12</b>	±11.44	/	/	/	/