	文件类型	技术文件	文件版本	V11.11.1	
	项目代号	/	页码	第0页 共4页	
	文件编号	GL DM-56-02-en			
文件名称： <p style="text-align: center;">CALIBRATOR NK1125 2025-12-10</p> 适用范围： 公司所有五分类仪器					
编制：刘金盈		审核：冯会娟		批准：程军	
日期：2025.11.04		日期：2025.11.04		日期：2025.11.04	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.11.0	/	新建	刘金盈	程军	2025.10.23
V11.11.1	/	增加1107A6.0至6.5版本参考值	刘金盈	程军	2025.11.04
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT NK1125

M 2025-10-12

H 2025-12-10

Applicable Instruments	Parameter	Unit	Reference Value
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11 or higher)	WBC	$\times 10^9/L$	9.68
	RBC	$\times 10^{12}/L$	4.21
	HGB	g/L	123
	MCV	fL	87.9
	PLT	$\times 10^9/L$	222
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.36
	RBC	$\times 10^{12}/L$	4.35
	HGB	g/L	128
	MCV	fL	89.6
	PLT	$\times 10^9/L$	223


【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 NK1125

 2025-10-12

 2025-12-10

Applicable Instruments	Parameter	Unit	Reference Value
DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	127
	MCV	fL	89.6
	PLT	$\times 10^9/L$	209
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	8.95
	RBC	$\times 10^{12}/L$	4.28
	HGB	g/L	125
	MCV	fL	85.4
	PLT	$\times 10^9/L$	233
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	8.83
	RBC	$\times 10^{12}/L$	4.19
	HGB	g/L	127
	MCV	fL	86.2
	PLT	$\times 10^9/L$	218
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.10
	HGB	g/L	125
	MCV	fL	89.0
	PLT	$\times 10^9/L$	220


【NOTE】


- The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT NK1125

 2025-10-12

 2025-12-10



Applicable Instruments	Parameter	Unit	Reference Value
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	9.49
	RBC	$\times 10^{12}/L$	4.38
	HGB	g/L	127
	MCV	fL	87.0
	PLT	$\times 10^9/L$	223
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.70
	RBC	$\times 10^{12}/L$	4.31
	HGB	g/L	128
	MCV	fL	86.6
	PLT	$\times 10^9/L$	228
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	9.88
	RBC	$\times 10^{12}/L$	4.37
	HGB	g/L	129
	MCV	fL	86.1
	PLT	$\times 10^9/L$	220
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	9.69
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	128
	MCV	fL	88.0
	PLT	$\times 10^9/L$	233

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK1125**
 **2025-10-12**
 **2025-12-10**

Applicable Instruments	Parameter	Unit	Reference Value
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	8.91
	RBC	$\times 10^{12}/L$	4.28
	HGB	g/L	124
	MCV	fL	86.1
	PLT	$\times 10^9/L$	221
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	9.81
	RBC	$\times 10^{12}/L$	4.19
	HGB	g/L	127
	MCV	fL	86.2
	PLT	$\times 10^9/L$	236
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.26
	RBC	$\times 10^{12}/L$	4.31
	HGB	g/L	127
	MCV	fL	86.1
	PLT	$\times 10^9/L$	220
DM60 Vet, DM61 Vet DM62 Vet, DM63 Vet (Technical File Version A1.4 or higher)	WBC	$\times 10^9/L$	8.12
	RBC	$\times 10^{12}/L$	3.86
	HGB	g/L	120
	MCV	fL	87.1
	PLT	$\times 10^9/L$	224

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.