



	文件类型	技术文件	文件版本	V5.11.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;"><b>CONTROL DM-RET RET1125 2026-01-25</b></p> 适用范围： 1116项目、7003项目						
编制：刘金盈 日期：2025.11.13		审核：冯会娟 日期：2025.11.13		批准：程军 日期：2025.11.13		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.11.0	/	新建		刘金盈	程军	2025.11.13
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库  <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-10-28

 2026-01-25

Applicable Instruments	Parameter	Unit	LOT RET1125L	LOT RET1125N	LOT RET1125H
DYMIND	RET#	$\times 10^{12}/L$	0.0452 $\pm 0.0226$	0.1763 $\pm 0.1058$	0.2030 $\pm 0.1015$
DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET%	%	0.80 $\pm 0.40$	3.98 $\pm 1.59$	8.98 $\pm 3.59$
	LFR	%	90.0 $\pm 27.0$	89.9 $\pm 27.0$	89.5 $\pm 27.0$
	MFR	%	8.0 $\pm 12.0$	8.0 $\pm 12.0$	8.6 $\pm 13.0$
	HFR	%	2.0 $\pm 3.0$	2.1 $\pm 3.0$	1.9 $\pm 3.0$
	IRF	%	10.0 $\pm 15.0$	10.1 $\pm 15.0$	10.5 $\pm 16.0$
	RHE	pg	15.9 $\pm 6.0$	15.2 $\pm 6.0$	15.5 $\pm 6.0$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET1125L



RET1125N




RET1125H


# DM-RET

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**

 2025-10-28

 2026-01-25

Applicable Instruments	Parameter	Unit	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET1125L	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET1125N	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET1125H
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0279 $\pm 0.0223$	0.1240 $\pm 0.0620$	0.1250 $\pm 0.0625$
	RET%	%	0.54 $\pm 0.43$	3.06 $\pm 1.53$	5.93 $\pm 2.97$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET1125L



RET1125N



RET1125H