


|  | 文件类型 | 技术文件 | 文件版本 | V3.12.0 | |
|---|---|---------------------------|------|--------------------------|------------|
| | 项目代号 | / | 页码 | 第0页, 共1页 | |
| | 文件编号 | GL DM-56-61-en | | | |
| 文件名称: <p style="text-align: center;">CALIBRATOR CPF1225 2026-01-10</p> 适用范围: 1117项目 | | | | | |
| 编制: 刘金盈 日期: 2025.11.20 | | 审核: 冯会娟 日期: 2025.11.20 | | 批准: 程军 日期: 2025.11.20 | |
| 修 订 记 录 | | | | | |
| 版本 | TCN/ECR/PCN编号 | 修订内容概要 | 修改人 | 批准人 | 批准日期 |
| V3.12.0 | / | 新建 | 刘金盈 | 程军 | 2025.11.20 |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| 发至: | <input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检 | | | | |
| 由工程 | | | | | |

DM-CAL-PF

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT CPF1225

2025-11-12

2026-01-10

| Applicable Instruments | Parameter | Unit | Reference Value |
|---|-----------|-----------------|-----------------|
| DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher) | PLT-F | $\times 10^9/L$ | 206 |

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.