


Table of Serum Amyloid A Protein Calibrator (SAA) Reference Value

Serum Amyloid A Protein Calibrator

LOT 2025112001

 2026-11-19

Applicable Instruments	DH-2000T series DH-100T、DH-102T DH-103T、DH-105T DH-106T、DH-108T (Technical File version A2.0 or higher)	DH-800 CS series (Technical File Version A1.3 or higher)
Level (multi-point)	Reference Value (mg/L)	Reference Value (mg/L)
Calibrator 1	10.73	11.25
Calibrator 2	23.01	25.55
Calibrator 3	51.52	58.78
Calibrator 4	110.79	123.37
Calibrator 5	170.33	184.51
Calibrator 6	196.42	215.64
Calibrator 7	227.13	260.48