



	文件类型	技术文件	文件版本	V5.6.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-18-en			
文件名称： <p style="text-align: center;">CONTROL DM-6D 6D1225 2026-02-10</p> 适用范围： 1116项目、1117项目					
编制：刘金盈 日期：2025.11.28		审核：冯会娟 日期：2025.11.28		批准：程军 日期：2025.11.28	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V5.6.0	/	新建	刘金盈	程军	2025.11.28
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND


CONTROL
 2025-11-13

 2026-02-10


Applicable Instruments	Parameter	Unit	LOT 6D1225L	LOT 6D1225N	LOT 6D1225H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	WBC	$\times 10^9/L$	3.99 ± 1.20	7.99 ± 1.80	21.90 ± 3.00
	Neu#	$\times 10^9/L$	2.95 ± 0.40	5.47 ± 0.80	16.34 ± 2.00
	Lym#	$\times 10^9/L$	0.52 ± 0.20	1.75 ± 0.50	4.36 ± 1.40
	Mon#	$\times 10^9/L$	0.39 ± 0.20	0.55 ± 0.40	0.85 ± 0.80
	Eos#	$\times 10^9/L$	0.13 ± 0.13	0.22 ± 0.22	0.35 ± 0.35
	Bas#	$\times 10^9/L$	3.34 ± 0.40	6.51 ± 0.70	16.45 ± 1.60
	IG#	$\times 10^9/L$	0.57 ± 0.20	1.04 ± 0.30	3.00 ± 0.80
	Neu%	%	73.9 ± 8.0	68.5 ± 7.0	74.6 ± 8.0
	Lym%	%	13.1 ± 5.0	21.9 ± 7.0	19.9 ± 6.0
	Mon%	%	9.7 ± 6.0	6.9 ± 5.0	3.9 ± 3.9
	Eos%	%	3.3 ± 3.3	2.7 ± 2.7	1.6 ± 1.6
	Bas%	%	83.8 ± 10.0	81.5 ± 9.0	75.1 ± 9.0
	IG%	%	14.2 ± 4.5	13.0 ± 4.7	13.7 ± 4.5
	RBC	$\times 10^{12}/L$	2.23 ± 0.20	4.47 ± 0.30	5.33 ± 0.50
	HGB	g/L	62 ± 4	135 ± 6	174 ± 8
	HCT	%	18.0 ± 2.0	38.9 ± 2.5	49.6 ± 3.0
	MCV	fL	80.5 ± 5.0	87.1 ± 5.0	93.1 ± 5.0
	MCH	pg	27.8 ± 2.5	30.2 ± 2.5	32.6 ± 2.5
	MCHC	g/L	344 ± 30	347 ± 30	351 ± 30
	RDW-CV	%	16.6 ± 5.0	15.4 ± 5.0	14.8 ± 6.0
	RDW-SD	fL	48.7 ± 10.0	49.3 ± 10.0	51.4 ± 12.0
	PLT	$\times 10^9/L$	55 ± 20	254 ± 45	524 ± 65
	MPV	fL	7.6 ± 3.0	7.3 ± 3.0	7.2 ± 3.0
	PDW	fL	9.8 ± 5.0	9.7 ± 5.0	9.8 ± 5.0
	PCT	%	0.042 ± 0.042	0.185 ± 0.100	0.377 ± 0.200
	P-LCR	%	13.9 ± 13.9	11.0 ± 11.0	11.1 ± 11.1
	P-LCC	$\times 10^9/L$	8 ± 8	28 ± 28	58 ± 50
IPF	%	3.1 ± 3.1	2.0 ± 2.0	1.7 ± 1.7	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D1225L



6D1225N





6D1225H

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND


CONTROL
 2025-11-13

 2026-02-10


Applicable Instruments	Parameter	Unit	LOT 6D1225L	LOT 6D1225N	LOT 6D1225H
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	WBC	$\times 10^9/L$	3.65 ± 1.20	7.78 ± 1.80	20.95 ± 3.00
	Neu#	$\times 10^9/L$	2.56 ± 0.40	5.01 ± 0.80	14.91 ± 2.00
	Lym#	$\times 10^9/L$	0.49 ± 0.20	1.72 ± 0.50	4.02 ± 1.40
	Mon#	$\times 10^9/L$	0.32 ± 0.20	0.48 ± 0.40	0.78 ± 0.78
	Eos#	$\times 10^9/L$	0.20 ± 0.20	0.40 ± 0.40	0.80 ± 0.80
	Bas#	$\times 10^9/L$	0.08 ± 0.08	0.17 ± 0.17	0.44 ± 0.44
	IG#	$\times 10^9/L$	0.50 ± 0.20	0.97 ± 0.30	2.89 ± 0.80
	Neu%	%	69.9 ± 8.0	64.4 ± 7.0	71.2 ± 8.0
	Lym%	%	13.3 ± 5.0	22.1 ± 7.0	19.2 ± 6.0
	Mon%	%	8.9 ± 6.0	6.2 ± 5.0	3.7 ± 3.7
	Eos%	%	5.6 ± 4.0	5.1 ± 4.0	3.8 ± 3.8
	Bas%	%	2.3 ± 2.3	2.2 ± 2.2	2.1 ± 2.1
	IG%	%	13.8 ± 4.5	12.5 ± 4.7	13.8 ± 4.5
	RBC	$\times 10^{12}/L$	2.36 ± 0.20	4.68 ± 0.30	5.63 ± 0.50
	HGB	g/L	64 ± 4	137 ± 6	178 ± 8
	HCT	%	19.7 ± 2.0	41.8 ± 2.5	54.0 ± 3.0
	MCV	fL	83.3 ± 5.0	89.3 ± 5.0	95.9 ± 5.0
	MCH	pg	27.1 ± 2.5	29.3 ± 2.5	31.6 ± 2.5
	MCHC	g/L	325 ± 30	328 ± 30	330 ± 30
	RDW-CV	%	17.6 ± 5.0	16.1 ± 5.0	15.4 ± 6.0
	RDW-SD	fL	51.3 ± 10.0	51.3 ± 10.0	53.1 ± 12.0
	PLT	$\times 10^9/L$	61 ± 20	246 ± 45	507 ± 65
	MPV	fL	9.1 ± 3.0	8.3 ± 3.0	8.3 ± 3.0
	PDW	fL	9.0 ± 5.0	8.4 ± 5.0	8.9 ± 5.0
	PCT	%	0.056 ± 0.050	0.204 ± 0.100	0.421 ± 0.200
	P-LCR	%	18.1 ± 18.1	13.7 ± 13.7	14.2 ± 14.2
	P-LCC	$\times 10^9/L$	11 ± 10	34 ± 30	72 ± 50
	IPF	%	12.5 ± 5.0	12.5 ± 5.0	12.5 ± 5.0
	NRBC#	$\times 10^9/L$	0.19 ± 0.19	0.01 ± 0.01	1.38 ± 0.40
	NRBC%	%	5.10 ± 4.00	0.10 ± 0.10	6.60 ± 4.00

【NOTE】


- The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D1225L



6D1225N



6D1225H