

DM-CAL-PF

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT CPF0126

2025-12-13

2026-02-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	PLT-F	$\times 10^9/L$	203

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.