



Coagulation Control Reference Value Table

Coagulation Control-level 2

CONTROL

LOT 2025090301

 20250903

 20270902

Applicable Instrument	Parameter	Unit	Principle	Reference Value	Deviation Limit
CA1200 CA1203 CA1205 DC1210 DC1250 DC1280	PT (Lyophilized)	s	Optical method	32.1	± 9.6
		s	Magnetic bead	32.6	± 9.8
	PT (liquid)	s	Optical method	31.5	± 9.4
		s	Magnetic bead	31.1	± 9.3
	APTT (Lyophilized)	s	Optical method	45.2	± 9.0
		s	Magnetic bead	44.0	± 8.8
	APTT (liquid)	s	Optical method	52.8	± 10.6
		s	Magnetic bead	49.6	± 9.9
	FIB (Lyophilized)	g/L	Optical method	1.48	± 0.44
		g/L	Magnetic bead	1.46	± 0.44
	FIB (liquid)	g/L	Optical method	1.50	± 0.45
		g/L	Magnetic bead	1.55	± 0.47
	AT-III (liquid)	%	Optical method	65.7	± 13.1

【NOTE】

1. The control should be stored in the refrigerator at 2 °C - 8 °C. After reconstitution, the control can be stable for 4 hours at room temperature (10 °C - 30 °C), and can be stable for 24 hours at 2 °C - 8 °C.
2. The control cannot be frozen after reconstitution.
3. The control is for IVD only.