



Coagulation Control Reference Value Table

Coagulation Control-level 1

CONTROL

LOT 2025101601

 20251016

 20271015

Applicable Instrument	Parameter	Unit	Principle	Reference Value	Deviation Limit
CA1200 CA1203 CA1205 DC1210 DC1250 DC1280	PT (Lyophilized)	s	Optical method	12.0	± 2.4
		s	Magnetic bead	12.3	± 2.5
	PT (liquid)	s	Optical method	12.4	± 2.5
		s	Magnetic bead	12.2	± 2.4
	APTT (Lyophilized)	s	Optical method	29.2	± 5.8
		s	Magnetic bead	28.3	± 5.7
	APTT (liquid)	s	Optical method	32.3	± 6.5
		s	Magnetic bead	30.2	± 6.0
	TT (Lyophilized)	s	Optical method	17.2	± 3.4
		s	Magnetic bead	18.5	± 3.7
	TT (liquid)	s	Optical method	20.9	± 4.2
		s	Magnetic bead	20.6	± 4.1
	FIB (Lyophilized)	g/L	Optical method	2.76	± 0.55
		g/L	Magnetic bead	2.82	± 0.56
	FIB (liquid)	g/L	Optical method	2.89	± 0.58
g/L		Magnetic bead	2.99	± 0.60	
AT-III (liquid)	%	Optical method	78.2	± 15.6	

【NOTE】

1. The control should be stored in the refrigerator at 2 °C - 8 °C. After reconstitution, the control can be stable for 4 hours at room temperature (10 °C - 30 °C), and can be stable for 24 hours at 2 °C - 8 °C.
2. The control cannot be frozen after reconstitution.
3. The control is for IVD only.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park,Guangqiao road,Tianliao Community,Yutangstreet,Guangming District, Shenzhen, Guangdong, P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132