



DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (Technical File Version A11.11 or higher)	WBC	$\times 10^9/L$	3.70 ± 0.50	9.03 ± 1.00	21.78 ± 2.50
	Neu%	%	47.2 ± 9.0	57.2 ± 8.0	65.4 ± 7.0
	Lym%	%	41.2 ± 9.0	30.3 ± 8.0	21.8 ± 6.0
	Mon%	%	7.6 ± 4.0	7.6 ± 5.0	7.8 ± 6.0
	Eos%	%	4.0 ± 4.0	4.9 ± 4.9	5.0 ± 5.0
	Bas%	%	60.2 ± 8.0	71.4 ± 8.0	79.7 ± 8.0
	Neu#	$\times 10^9/L$	1.75 ± 0.40	5.16 ± 0.70	14.24 ± 1.40
	Lym#	$\times 10^9/L$	1.52 ± 0.40	2.74 ± 0.70	4.75 ± 1.10
	Mon#	$\times 10^9/L$	0.28 ± 0.14	0.69 ± 0.50	1.70 ± 1.10
	Eos#	$\times 10^9/L$	0.15 ± 0.15	0.44 ± 0.44	1.09 ± 1.09
	Bas#	$\times 10^9/L$	2.23 ± 0.30	6.45 ± 0.70	17.36 ± 1.50
	RBC	$\times 10^{12}/L$	2.24 ± 0.18	4.60 ± 0.24	5.41 ± 0.50
	HGB	g/L	60 ± 4	134 ± 6	169 ± 8
	HCT	%	19.2 ± 2.0	41.2 ± 3.0	51.3 ± 4.0
	MCV	fL	85.8 ± 5.0	89.6 ± 5.0	94.8 ± 6.0
	MCH	pg	26.8 ± 2.5	29.1 ± 2.5	31.2 ± 2.5
	MCHC	g/L	312 ± 30	325 ± 30	330 ± 30
	RDW-CV	%	18.5 ± 3.0	16.8 ± 3.0	16.4 ± 3.0
	RDW-SD	fL	57.7 ± 10.0	55.3 ± 10.0	57.1 ± 12.0
	PLT	$\times 10^9/L$	53 ± 20	248 ± 40	498 ± 60
MPV	fL	9.9 ± 3.0	9.1 ± 3.0	9.0 ± 3.0	
PDW	fL	9.8 ± 3.0	10.3 ± 3.0	10.6 ± 3.0	
PCT	%	0.052 ± 0.050	0.226 ± 0.100	0.448 ± 0.200	
P-LCR	%	22.6 ± 8.0	20.6 ± 8.0	20.1 ± 8.0	
P-LCC	$\times 10^9/L$	12 ± 12	51 ± 25	100 ± 35	

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND UN71 UN73 UN76 DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	3.55 ± 0.50	8.55 ± 1.00	20.71 ± 2.50
	Neu%	%	48.5 ± 9.0	56.7 ± 8.0	63.5 ± 7.0
	Lym%	%	39.7 ± 9.0	29.4 ± 8.0	20.9 ± 6.0
	Mon%	%	6.3 ± 4.0	7.5 ± 5.0	7.3 ± 6.0
	Eos%	%	4.8 ± 4.8	5.6 ± 5.6	7.4 ± 7.0
	Bas%	%	0.7 ± 0.7	0.8 ± 0.8	0.9 ± 0.9
	Neu#	$\times 10^9/L$	1.73 ± 0.40	4.85 ± 0.70	13.15 ± 1.40
	Lym#	$\times 10^9/L$	1.41 ± 0.40	2.51 ± 0.70	4.33 ± 1.10
	Mon#	$\times 10^9/L$	0.22 ± 0.14	0.64 ± 0.50	1.51 ± 1.10
	Eos#	$\times 10^9/L$	0.17 ± 0.15	0.48 ± 0.48	1.53 ± 1.30
	Bas#	$\times 10^9/L$	0.02 ± 0.02	0.07 ± 0.07	0.19 ± 0.19
	RBC	$\times 10^{12}/L$	2.30 ± 0.18	4.69 ± 0.24	5.55 ± 0.50
	HGB	g/L	63 ± 4	138 ± 6	172 ± 8
	HCT	%	19.6 ± 2.0	42.4 ± 3.0	53.3 ± 4.0
	MCV	fL	85.0 ± 5.0	90.5 ± 5.0	96.0 ± 6.0
	MCH	pg	27.4 ± 2.5	29.4 ± 2.5	31.0 ± 2.5
	MCHC	g/L	322 ± 30	325 ± 30	323 ± 30
	RDW-CV	%	18.7 ± 3.0	17.1 ± 3.0	16.6 ± 3.0
	RDW-SD	fL	59.6 ± 10.0	57.4 ± 10.0	59.2 ± 12.0
	PLT	$\times 10^9/L$	52 ± 20	242 ± 40	483 ± 60
MPV	fL	10.0 ± 3.0	9.1 ± 3.0	9.0 ± 3.0	
PDW	fL	10.3 ± 3.0	10.5 ± 3.0	10.3 ± 3.0	
PCT	%	0.052 ± 0.050	0.220 ± 0.100	0.435 ± 0.200	
P-LCR	%	26.9 ± 8.0	20.7 ± 8.0	20.3 ± 8.0	
P-LCC	$\times 10^9/L$	14 ± 14	50 ± 25	98 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	2.37 ± 0.18	4.83 ± 0.24	5.64 ± 0.50
	HGB	g/L	62 ± 4	136 ± 6	171 ± 8
	HCT	%	20.4 ± 2.0	44.1 ± 3.0	54.9 ± 4.0
	MCV	fL	85.9 ± 5.0	91.4 ± 5.0	97.3 ± 6.0
	MCH	pg	26.2 ± 2.5	28.2 ± 2.5	30.3 ± 2.5
	MCHC	g/L	305 ± 30	308 ± 30	312 ± 30
	RDW-CV	%	18.6 ± 3.0	17.0 ± 3.0	16.7 ± 3.0
	RDW-SD	fL	55.3 ± 10.0	53.0 ± 10.0	54.9 ± 12.0
	PLT	$\times 10^9/L$	54 ± 20	229 ± 40	445 ± 60
	MPV	fL	9.2 ± 3.0	9.1 ± 3.0	9.1 ± 3.0
	PDW	fL	11.2 ± 3.0	10.5 ± 3.0	10.4 ± 3.0
	PCT	%	0.050 ± 0.050	0.208 ± 0.100	0.405 ± 0.200
	P-LCR	%	22.2 ± 8.0	21.0 ± 8.0	20.7 ± 8.0
	P-LCC	$\times 10^9/L$	12 ± 12	48 ± 25	92 ± 35

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	3.61 ± 0.50	8.91 ± 1.00	21.16 ± 2.50
	Neu%	%	48.7 ± 9.0	57.2 ± 8.0	64.4 ± 7.0
	Lym%	%	42.0 ± 9.0	32.5 ± 8.0	22.4 ± 6.0
	Mon%	%	4.4 ± 4.0	3.9 ± 3.9	4.6 ± 4.6
	Eos%	%	4.9 ± 4.9	6.4 ± 6.0	8.6 ± 7.0
	Bas%	%	1.6 ± 1.6	3.3 ± 3.3	2.7 ± 2.7
	Neu#	$\times 10^9/L$	1.75 ± 0.40	5.09 ± 0.70	13.63 ± 1.40
	Lym#	$\times 10^9/L$	1.52 ± 0.40	2.90 ± 0.70	4.74 ± 1.10
	Mon#	$\times 10^9/L$	0.16 ± 0.14	0.35 ± 0.35	0.97 ± 0.97
	Eos#	$\times 10^9/L$	0.18 ± 0.15	0.57 ± 0.50	1.82 ± 1.30
	Bas#	$\times 10^9/L$	0.06 ± 0.06	0.29 ± 0.29	0.57 ± 0.57
	RBC	$\times 10^{12}/L$	2.28 ± 0.18	4.65 ± 0.24	5.42 ± 0.50
	HGB	g/L	61 ± 4	136 ± 6	172 ± 8
	HCT	%	19.3 ± 2.0	40.8 ± 3.0	50.1 ± 4.0
	MCV	fL	84.7 ± 5.0	87.8 ± 5.0	92.4 ± 6.0
	MCH	pg	26.8 ± 2.5	29.2 ± 2.5	31.7 ± 2.5
	MCHC	g/L	316 ± 30	333 ± 30	343 ± 30
	RDW-CV	%	15.6 ± 3.0	14.6 ± 3.0	14.4 ± 3.0
	RDW-SD	fL	55.7 ± 10.0	54.4 ± 10.0	56.5 ± 12.0
	PLT	$\times 10^9/L$	65 ± 20	246 ± 40	453 ± 60
MPV	fL	9.3 ± 3.0	8.9 ± 3.0	9.0 ± 3.0	
PDW	fL	11.0 ± 3.0	11.7 ± 3.0	11.8 ± 3.0	
PCT	%	0.060 ± 0.050	0.219 ± 0.100	0.408 ± 0.200	
P-LCR	%	29.2 ± 8.0	28.9 ± 8.0	29.6 ± 8.0	
P-LCC	$\times 10^9/L$	19 ± 15	71 ± 25	134 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



BC0126L




BC0126N




BC0126H

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	3.55 ± 0.50	8.81 ± 1.00	20.91 ± 2.50
	Neu%	%	49.0 ± 9.0	57.7 ± 8.0	64.9 ± 7.0
	Lym%	%	41.3 ± 9.0	31.2 ± 8.0	21.8 ± 6.0
	Mon%	%	4.8 ± 4.0	4.5 ± 4.5	4.6 ± 4.6
	Eos%	%	4.9 ± 4.9	6.6 ± 6.0	8.7 ± 7.0
	Bas%	%	1.7 ± 1.7	3.1 ± 3.1	2.4 ± 2.4
	Neu#	$\times 10^9/L$	1.74 ± 0.40	5.08 ± 0.70	13.57 ± 1.40
	Lym#	$\times 10^9/L$	1.47 ± 0.40	2.75 ± 0.70	4.56 ± 1.10
	Mon#	$\times 10^9/L$	0.17 ± 0.14	0.40 ± 0.40	0.96 ± 0.96
	Eos#	$\times 10^9/L$	0.17 ± 0.15	0.58 ± 0.50	1.82 ± 1.30
	Bas#	$\times 10^9/L$	0.06 ± 0.06	0.27 ± 0.27	0.50 ± 0.50
	RBC	$\times 10^{12}/L$	2.20 ± 0.18	4.44 ± 0.24	5.17 ± 0.50
	HGB	g/L	60 ± 4	135 ± 6	171 ± 8
	HCT	%	18.5 ± 2.0	39.1 ± 3.0	47.9 ± 4.0
	MCV	fL	84.2 ± 5.0	88.1 ± 5.0	92.7 ± 6.0
	MCH	pg	27.3 ± 2.5	30.4 ± 2.5	33.1 ± 2.5
	MCHC	g/L	324 ± 30	345 ± 30	357 ± 30
	RDW-CV	%	15.6 ± 3.0	14.6 ± 3.0	14.2 ± 3.0
	RDW-SD	fL	54.2 ± 10.0	53.1 ± 10.0	54.6 ± 12.0
	PLT	$\times 10^9/L$	61 ± 20	243 ± 40	454 ± 60
MPV	fL	8.8 ± 3.0	8.3 ± 3.0	8.5 ± 3.0	
PDW	fL	10.4 ± 3.0	11.6 ± 3.0	11.7 ± 3.0	
PCT	%	0.054 ± 0.050	0.202 ± 0.100	0.386 ± 0.200	
P-LCR	%	24.6 ± 8.0	24.7 ± 8.0	25.3 ± 8.0	
P-LCC	$\times 10^9/L$	15 ± 15	60 ± 25	115 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not applicable to DF50,DF51,DF53, DF50 Vet and DF52 Vet Instruments.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DH71 DH76 (Technical File Version A4.1to A10.3) DH51 DH53 DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	3.59 ± 0.50	8.40 ± 1.00	19.99 ± 2.50
	Neu%	%	48.1 ± 9.0	59.2 ± 8.0	65.4 ± 7.0
	Lym%	%	41.4 ± 9.0	29.8 ± 8.0	21.5 ± 6.0
	Mon%	%	6.2 ± 4.0	6.3 ± 5.0	6.8 ± 6.0
	Eos%	%	4.3 ± 4.3	4.7 ± 4.7	6.3 ± 6.3
	Bas%	%	59.8 ± 8.0	70.6 ± 8.0	79.4 ± 8.0
	Neu#	$\times 10^9/L$	1.73 ± 0.40	4.98 ± 0.70	13.07 ± 1.40
	Lym#	$\times 10^9/L$	1.49 ± 0.40	2.50 ± 0.70	4.30 ± 1.10
	Mon#	$\times 10^9/L$	0.22 ± 0.14	0.53 ± 0.50	1.36 ± 1.10
	Eos#	$\times 10^9/L$	0.15 ± 0.15	0.39 ± 0.39	1.26 ± 1.26
	Bas#	$\times 10^9/L$	2.15 ± 0.30	5.93 ± 0.70	15.87 ± 1.50
	RBC	$\times 10^{12}/L$	2.19 ± 0.18	4.53 ± 0.24	5.35 ± 0.50
	HGB	g/L	61 ± 4	136 ± 6	173 ± 8
	HCT	%	18.9 ± 2.0	41.5 ± 3.0	52.3 ± 4.0
	MCV	fL	86.1 ± 5.0	91.6 ± 5.0	97.7 ± 6.0
	MCH	pg	27.9 ± 2.5	30.0 ± 2.5	32.3 ± 2.5
	MCHC	g/L	324 ± 30	328 ± 30	331 ± 30
	RDW-CV	%	18.4 ± 3.0	17.0 ± 3.0	16.4 ± 3.0
	RDW-SD	fL	55.0 ± 10.0	53.3 ± 10.0	54.8 ± 12.0
	PLT	$\times 10^9/L$	51 ± 20	239 ± 40	468 ± 60
MPV	fL	10.6 ± 3.0	9.5 ± 3.0	9.5 ± 3.0	
PDW	fL	11.5 ± 3.0	11.4 ± 3.0	11.3 ± 3.0	
PCT	%	0.054 ± 0.050	0.227 ± 0.100	0.445 ± 0.200	
P-LCR	%	27.5 ± 8.0	23.4 ± 8.0	22.9 ± 8.0	
P-LCC	$\times 10^9/L$	14 ± 14	56 ± 25	107 ± 35	

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DH71 DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	3.69 ± 0.50	8.84 ± 1.00	20.74 ± 2.50
	Neu%	%	48.6 ± 9.0	57.8 ± 8.0	64.6 ± 7.0
	Lym%	%	40.6 ± 9.0	30.7 ± 8.0	22.2 ± 6.0
	Mon%	%	7.0 ± 4.0	7.3 ± 5.0	7.2 ± 6.0
	Eos%	%	3.8 ± 3.8	4.2 ± 4.2	6.0 ± 6.0
	Bas%	%	60.4 ± 8.0	71.2 ± 8.0	80.1 ± 8.0
	Neu#	$\times 10^9/L$	1.79 ± 0.40	5.11 ± 0.70	13.41 ± 1.40
	Lym#	$\times 10^9/L$	1.50 ± 0.40	2.71 ± 0.70	4.60 ± 1.10
	Mon#	$\times 10^9/L$	0.26 ± 0.14	0.65 ± 0.50	1.49 ± 1.10
	Eos#	$\times 10^9/L$	0.14 ± 0.14	0.37 ± 0.37	1.24 ± 1.24
	Bas#	$\times 10^9/L$	2.23 ± 0.30	6.29 ± 0.70	16.61 ± 1.50
	RBC	$\times 10^{12}/L$	2.30 ± 0.18	4.75 ± 0.24	5.56 ± 0.50
	HGB	g/L	61 ± 4	136 ± 6	169 ± 8
	HCT	%	19.2 ± 2.0	41.8 ± 3.0	51.8 ± 4.0
	MCV	fL	83.5 ± 5.0	87.9 ± 5.0	93.1 ± 6.0
	MCH	pg	26.5 ± 2.5	28.6 ± 2.5	30.4 ± 2.5
	MCHC	g/L	318 ± 30	326 ± 30	326 ± 30
	RDW-CV	%	18.5 ± 3.0	16.9 ± 3.0	16.5 ± 3.0
	RDW-SD	fL	56.0 ± 10.0	53.4 ± 10.0	55.4 ± 12.0
	PLT	$\times 10^9/L$	52 ± 20	244 ± 40	484 ± 60
MPV	fL	9.8 ± 3.0	9.0 ± 3.0	8.9 ± 3.0	
PDW	fL	10.0 ± 3.0	10.3 ± 3.0	10.3 ± 3.0	
PCT	%	0.051 ± 0.050	0.220 ± 0.100	0.431 ± 0.200	
P-LCR	%	23.1 ± 8.0	20.1 ± 8.0	19.6 ± 8.0	
P-LCC	$\times 10^9/L$	12 ± 12	49 ± 25	95 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (Technical File Version A6.0 to A 6.5)	WBC	$\times 10^9/L$	3.67 ± 0.50	8.60 ± 1.00	20.42 ± 2.50
	Neu%	%	46.8 ± 9.0	56.7 ± 8.0	63.5 ± 7.0
	Lym%	%	41.6 ± 9.0	30.5 ± 8.0	21.9 ± 6.0
	Mon%	%	7.0 ± 4.0	7.3 ± 5.0	7.0 ± 6.0
	Eos%	%	4.6 ± 4.6	5.5 ± 5.5	7.6 ± 7.0
	Bas%	%	60.9 ± 8.0	71.4 ± 8.0	79.7 ± 8.0
	Neu#	$\times 10^9/L$	1.71 ± 0.40	4.88 ± 0.70	12.97 ± 1.40
	Lym#	$\times 10^9/L$	1.53 ± 0.40	2.62 ± 0.70	4.47 ± 1.10
	Mon#	$\times 10^9/L$	0.26 ± 0.14	0.63 ± 0.50	1.43 ± 1.10
	Eos#	$\times 10^9/L$	0.17 ± 0.15	0.47 ± 0.47	1.55 ± 1.30
	Bas#	$\times 10^9/L$	2.24 ± 0.30	6.14 ± 0.70	16.27 ± 1.50
	RBC	$\times 10^{12}/L$	2.27 ± 0.18	4.66 ± 0.24	5.45 ± 0.50
	HGB	g/L	61 ± 4	136 ± 6	171 ± 8
	HCT	%	19.2 ± 2.0	41.2 ± 3.0	50.9 ± 4.0
	MCV	fL	84.5 ± 5.0	88.4 ± 5.0	93.4 ± 6.0
	MCH	pg	26.9 ± 2.5	29.2 ± 2.5	31.4 ± 2.5
	MCHC	g/L	318 ± 30	330 ± 30	336 ± 30
	RDW-CV	%	18.6 ± 3.0	17.1 ± 3.0	16.6 ± 3.0
	RDW-SD	fL	56.4 ± 10.0	54.3 ± 10.0	56.0 ± 12.0
	PLT	$\times 10^9/L$	53 ± 20	247 ± 40	503 ± 60
MPV	fL	10.3 ± 3.0	9.5 ± 3.0	9.5 ± 3.0	
PDW	fL	10.6 ± 3.0	11.4 ± 3.0	11.3 ± 3.0	
PCT	%	0.055 ± 0.050	0.235 ± 0.100	0.478 ± 0.200	
P-LCR	%	26.4 ± 8.0	23.9 ± 8.0	22.9 ± 8.0	
P-LCC	$\times 10^9/L$	14 ± 14	59 ± 25	115 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N





BC0126H

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	3.83 ± 0.50	9.12 ± 1.00	21.63 ± 2.50
	Neu%	%	47.4 ± 9.0	57.5 ± 8.0	63.8 ± 7.0
	Lym%	%	41.1 ± 9.0	30.6 ± 8.0	22.3 ± 6.0
	Mon%	%	6.8 ± 4.0	6.9 ± 5.0	7.0 ± 6.0
	Eos%	%	4.7 ± 4.7	5.0 ± 5.0	6.9 ± 6.9
	Bas%	%	61.5 ± 8.0	71.8 ± 8.0	80.5 ± 8.0
	Neu#	$\times 10^9/L$	1.82 ± 0.40	5.24 ± 0.70	13.81 ± 1.40
	Lym#	$\times 10^9/L$	1.57 ± 0.40	2.79 ± 0.70	4.82 ± 1.10
	Mon#	$\times 10^9/L$	0.26 ± 0.14	0.63 ± 0.50	1.51 ± 1.10
	Eos#	$\times 10^9/L$	0.18 ± 0.15	0.46 ± 0.46	1.49 ± 1.30
	Bas#	$\times 10^9/L$	2.36 ± 0.30	6.55 ± 0.70	17.41 ± 1.50
	RBC	$\times 10^{12}/L$	2.29 ± 0.18	4.67 ± 0.24	5.46 ± 0.50
	HGB	g/L	61 ± 4	137 ± 6	171 ± 8
	HCT	%	19.2 ± 2.0	41.0 ± 3.0	50.8 ± 4.0
	MCV	fL	83.7 ± 5.0	87.8 ± 5.0	93.0 ± 6.0
	MCH	pg	26.6 ± 2.5	29.3 ± 2.5	31.3 ± 2.5
	MCHC	g/L	318 ± 30	334 ± 30	337 ± 30
	RDW-CV	%	18.3 ± 3.0	16.7 ± 3.0	16.2 ± 3.0
	RDW-SD	fL	55.6 ± 10.0	53.3 ± 10.0	54.9 ± 12.0
	PLT	$\times 10^9/L$	52 ± 20	238 ± 40	490 ± 60
MPV	fL	9.5 ± 3.0	8.7 ± 3.0	8.7 ± 3.0	
PDW	fL	9.1 ± 3.0	9.9 ± 3.0	9.6 ± 3.0	
PCT	%	0.049 ± 0.049	0.207 ± 0.100	0.426 ± 0.200	
P-LCR	%	21.2 ± 8.0	18.5 ± 8.0	17.6 ± 8.0	
P-LCC	$\times 10^9/L$	11 ± 11	44 ± 25	86 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND D6-CRP DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	3.78 ± 0.50	8.90 ± 1.00	20.78 ± 2.50
	Neu%	%	47.1 ± 9.0	56.7 ± 8.0	64.1 ± 7.0
	Lym%	%	41.4 ± 9.0	31.0 ± 8.0	22.2 ± 6.0
	Mon%	%	7.1 ± 4.0	7.4 ± 5.0	7.5 ± 6.0
	Eos%	%	4.4 ± 4.4	4.9 ± 4.9	6.2 ± 6.2
	Bas%	%	60.9 ± 8.0	71.7 ± 8.0	80.4 ± 8.0
	Neu#	$\times 10^9/L$	1.78 ± 0.40	5.04 ± 0.70	13.32 ± 1.40
	Lym#	$\times 10^9/L$	1.56 ± 0.40	2.76 ± 0.70	4.61 ± 1.10
	Mon#	$\times 10^9/L$	0.27 ± 0.14	0.66 ± 0.50	1.56 ± 1.10
	Eos#	$\times 10^9/L$	0.17 ± 0.15	0.44 ± 0.44	1.29 ± 1.29
	Bas#	$\times 10^9/L$	2.30 ± 0.30	6.38 ± 0.70	16.71 ± 1.50
	RBC	$\times 10^{12}/L$	2.37 ± 0.18	4.78 ± 0.24	5.58 ± 0.50
	HGB	g/L	64 ± 4	139 ± 6	172 ± 8
	HCT	%	20.2 ± 2.0	42.8 ± 3.0	52.8 ± 4.0
	MCV	fL	85.4 ± 5.0	89.5 ± 5.0	94.7 ± 6.0
	MCH	pg	27.0 ± 2.5	29.1 ± 2.5	30.8 ± 2.5
	MCHC	g/L	316 ± 30	325 ± 30	325 ± 30
	RDW-CV	%	18.3 ± 3.0	16.8 ± 3.0	16.2 ± 3.0
	RDW-SD	fL	57.5 ± 10.0	55.2 ± 10.0	56.6 ± 12.0
	PLT	$\times 10^9/L$	55 ± 20	246 ± 40	494 ± 60
MPV	fL	9.3 ± 3.0	9.1 ± 3.0	9.0 ± 3.0	
PDW	fL	11.0 ± 3.0	10.5 ± 3.0	10.3 ± 3.0	
PCT	%	0.051 ± 0.050	0.224 ± 0.100	0.445 ± 0.200	
P-LCR	%	23.6 ± 8.0	20.7 ± 8.0	19.8 ± 8.0	
P-LCC	$\times 10^9/L$	13 ± 13	51 ± 25	98 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L




BC0126N




BC0126H

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DF50 Vet DF52 Vet DF55 Vet DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	×10 ⁹ /L	3.41 ±0.50	8.40 ±1.00	19.98 ±2.50
	Neu%	%	51.7 ±15.0	54.3 ±15.0	65.0 ±15.0
	Lym%	%	40.7 ±9.0	36.7 ±15.0	19.7 ±10.0
	Mon%	%	3.5 ±3.5	3.2 ±3.2	8.1 ±7.0
	Eos%	%	4.1 ±4.1	5.8 ±5.8	7.2 ±7.2
	Bas%	%	1.6 ±1.6	3.3 ±3.3	2.3 ±2.3
	Neu#	×10 ⁹ /L	1.76 ±0.60	4.56 ±1.40	12.98 ±2.80
	Lym#	×10 ⁹ /L	1.39 ±0.40	3.08 ±1.10	3.94 ±2.50
	Mon#	×10 ⁹ /L	0.12 ±0.12	0.27 ±0.27	1.62 ±1.62
	Eos#	×10 ⁹ /L	0.14 ±0.14	0.49 ±0.49	1.44 ±1.44
	Bas#	×10 ⁹ /L	0.05 ±0.05	0.28 ±0.28	0.46 ±0.46
	RBC	×10 ¹² /L	2.24 ±0.18	4.55 ±0.24	5.28 ±0.50
	HGB	g/L	59 ±4	133 ±6	169 ±8
	HCT	%	18.6 ±2.0	40.0 ±3.0	49.6 ±4.0
	MCV	fL	82.9 ±5.0	87.9 ±5.0	93.9 ±6.0
	MCH	pg	26.3 ±2.5	29.2 ±2.5	32.0 ±2.5
	MCHC	g/L	318 ±30	333 ±30	341 ±30
	RDW-CV	%	14.3 ±3.0	13.3 ±3.0	12.9 ±3.0
	RDW-SD	fL	50.1 ±10.0	48.5 ±10.0	49.9 ±12.0
	PLT	×10 ⁹ /L	62 ±20	232 ±40	424 ±60
MPV	fL	9.0 ±3.0	8.5 ±3.0	8.6 ±3.0	
PDW	fL	10.6 ±3.0	11.3 ±3.0	11.3 ±3.0	
PCT	%	0.056 ±0.050	0.197 ±0.100	0.365 ±0.200	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	3.73 ± 0.50	9.06 ± 1.00	21.89 ± 2.50
	Neu%	%	48.3 ± 9.0	57.3 ± 8.0	64.5 ± 7.0
	Lym%	%	43.0 ± 9.0	31.3 ± 8.0	21.9 ± 6.0
	Mon%	%	4.0 ± 4.0	4.9 ± 4.9	5.2 ± 5.2
	Eos%	%	4.7 ± 4.7	6.5 ± 6.0	8.4 ± 7.0
	Bas%	%	1.8 ± 1.8	3.7 ± 3.7	3.0 ± 3.0
	Neu#	$\times 10^9/L$	1.80 ± 0.40	5.19 ± 0.70	14.12 ± 1.40
	Lym#	$\times 10^9/L$	1.60 ± 0.40	2.84 ± 0.70	4.79 ± 1.10
	Mon#	$\times 10^9/L$	0.15 ± 0.14	0.44 ± 0.44	1.14 ± 1.10
	Eos#	$\times 10^9/L$	0.18 ± 0.15	0.59 ± 0.50	1.84 ± 1.30
	Bas#	$\times 10^9/L$	0.07 ± 0.07	0.34 ± 0.34	0.66 ± 0.66
	RBC	$\times 10^{12}/L$	2.23 ± 0.18	4.54 ± 0.24	5.29 ± 0.50
	HGB	g/L	62 ± 4	135 ± 6	170 ± 8
	HCT	%	18.9 ± 2.0	40.0 ± 3.0	49.1 ± 4.0
	MCV	fL	84.7 ± 5.0	88.0 ± 5.0	92.8 ± 6.0
	MCH	pg	27.8 ± 2.5	29.7 ± 2.5	32.1 ± 2.5
	MCHC	g/L	328 ± 30	338 ± 30	346 ± 30
	RDW-CV	%	16.3 ± 3.0	15.0 ± 3.0	14.5 ± 3.0
	RDW-SD	fL	58.6 ± 10.0	57.5 ± 10.0	59.3 ± 12.0
	PLT	$\times 10^9/L$	49 ± 20	247 ± 40	502 ± 60
MPV	fL	9.1 ± 3.0	8.3 ± 3.0	8.4 ± 3.0	
PDW	fL	9.9 ± 3.0	10.8 ± 3.0	10.8 ± 3.0	
PCT	%	0.045 ± 0.045	0.205 ± 0.100	0.422 ± 0.200	
P-LCR	%	26.5 ± 8.0	24.7 ± 8.0	25.1 ± 8.0	
P-LCC	$\times 10^9/L$	13 ± 13	61 ± 25	126 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DM71X DM72X DM78X DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	3.62 ± 0.50	8.47 ± 1.00	19.91 ± 2.50
	Neu%	%	47.9 ± 9.0	57.3 ± 8.0	64.4 ± 7.0
	Lym%	%	40.1 ± 9.0	29.9 ± 8.0	20.9 ± 6.0
	Mon%	%	7.5 ± 4.0	7.3 ± 5.0	7.4 ± 6.0
	Eos%	%	4.5 ± 4.5	5.5 ± 5.5	7.3 ± 7.0
	Bas%	%	60.4 ± 8.0	70.9 ± 8.0	80.1 ± 8.0
	Neu#	$\times 10^9/L$	1.74 ± 0.40	4.85 ± 0.70	12.83 ± 1.40
	Lym#	$\times 10^9/L$	1.45 ± 0.40	2.53 ± 0.70	4.16 ± 1.10
	Mon#	$\times 10^9/L$	0.27 ± 0.14	0.62 ± 0.50	1.47 ± 1.10
	Eos#	$\times 10^9/L$	0.16 ± 0.15	0.47 ± 0.47	1.45 ± 1.30
	Bas#	$\times 10^9/L$	2.19 ± 0.30	6.01 ± 0.70	15.95 ± 1.50
	RBC	$\times 10^{12}/L$	2.27 ± 0.18	4.68 ± 0.24	5.47 ± 0.50
	HGB	g/L	63 ± 4	136 ± 6	169 ± 8
	HCT	%	18.9 ± 2.0	41.4 ± 3.0	51.4 ± 4.0
	MCV	fL	83.2 ± 5.0	88.5 ± 5.0	93.9 ± 6.0
	MCH	pg	27.8 ± 2.5	29.1 ± 2.5	30.9 ± 2.5
	MCHC	g/L	334 ± 30	328 ± 30	329 ± 30
	RDW-CV	%	18.7 ± 3.0	17.1 ± 3.0	16.6 ± 3.0
	RDW-SD	fL	56.3 ± 10.0	54.2 ± 10.0	55.8 ± 12.0
	PLT	$\times 10^9/L$	51 ± 20	240 ± 40	488 ± 60
MPV	fL	9.0 ± 3.0	8.8 ± 3.0	8.8 ± 3.0	
PDW	fL	10.9 ± 3.0	10.0 ± 3.0	10.1 ± 3.0	
PCT	%	0.046 ± 0.046	0.211 ± 0.100	0.429 ± 0.200	
P-LCR	%	21.6 ± 8.0	18.8 ± 8.0	18.9 ± 8.0	
P-LCC	$\times 10^9/L$	11 ± 11	45 ± 25	92 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N




BC0126H


DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND UN71 Vet UN73 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	3.48 ± 0.50	8.36 ± 1.00	19.93 ± 2.50
	Neu%	%	48.4 ± 9.0	57.6 ± 8.0	64.6 ± 7.0
	Lym%	%	40.2 ± 9.0	29.6 ± 8.0	20.9 ± 6.0
	Mon%	%	6.7 ± 4.0	7.3 ± 5.0	7.3 ± 6.0
	Eos%	%	4.7 ± 4.7	5.5 ± 5.5	7.2 ± 7.0
	Bas%	%	0.7 ± 0.7	0.8 ± 0.8	0.9 ± 0.9
	Neu#	$\times 10^9/L$	1.69 ± 0.40	4.82 ± 0.70	12.88 ± 1.40
	Lym#	$\times 10^9/L$	1.40 ± 0.40	2.47 ± 0.70	4.17 ± 1.10
	Mon#	$\times 10^9/L$	0.23 ± 0.14	0.61 ± 0.50	1.45 ± 1.10
	Eos#	$\times 10^9/L$	0.16 ± 0.15	0.46 ± 0.46	1.43 ± 1.30
	Bas#	$\times 10^9/L$	0.02 ± 0.02	0.07 ± 0.07	0.18 ± 0.18
	RBC	$\times 10^{12}/L$	2.32 ± 0.18	4.73 ± 0.24	5.50 ± 0.50
	HGB	g/L	64 ± 4	137 ± 6	171 ± 8
	HCT	%	19.4 ± 2.0	42.1 ± 3.0	52.0 ± 4.0
	MCV	fL	83.8 ± 5.0	89.0 ± 5.0	94.6 ± 6.0
	MCH	pg	27.6 ± 2.5	29.0 ± 2.5	31.1 ± 2.5
	MCHC	g/L	329 ± 30	325 ± 30	329 ± 30
	RDW-CV	%	18.6 ± 3.0	17.0 ± 3.0	16.4 ± 3.0
	RDW-SD	fL	49.4 ± 10.0	47.1 ± 10.0	48.4 ± 12.0
	PLT	$\times 10^9/L$	53 ± 20	241 ± 40	465 ± 60
MPV	fL	9.4 ± 3.0	8.4 ± 3.0	8.4 ± 3.0	
PDW	fL	8.7 ± 3.0	9.3 ± 3.0	9.1 ± 3.0	
PCT	%	0.050 ± 0.050	0.202 ± 0.100	0.391 ± 0.200	
P-LCR	%	20.8 ± 8.0	17.0 ± 8.0	16.6 ± 8.0	
P-LCC	$\times 10^9/L$	11 ± 11	41 ± 25	77 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N





BC0126H

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL
 2025-12-11

 2026-03-10

Applicable Instruments	Parameter	Unit	LOT BC0126L	LOT BC0126N	LOT BC0126H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	WBC	$\times 10^9/L$	3.09 ± 0.50	7.67 ± 1.00	18.35 ± 2.50
	Neu%	%	49.4 ± 9.0	58.6 ± 8.0	65.9 ± 7.0
	Lym%	%	42.0 ± 9.0	31.0 ± 8.0	20.7 ± 6.0
	Mon%	%	4.0 ± 4.0	3.9 ± 3.9	5.0 ± 5.0
	Eos%	%	4.6 ± 4.6	6.5 ± 6.0	8.4 ± 7.0
	Bas%	%	1.9 ± 1.9	3.5 ± 3.5	2.8 ± 2.8
	Neu#	$\times 10^9/L$	1.53 ± 0.40	4.49 ± 0.70	12.09 ± 1.40
	Lym#	$\times 10^9/L$	1.30 ± 0.40	2.38 ± 0.70	3.80 ± 1.10
	Mon#	$\times 10^9/L$	0.12 ± 0.12	0.30 ± 0.30	0.92 ± 0.92
	Eos#	$\times 10^9/L$	0.14 ± 0.14	0.50 ± 0.50	1.54 ± 1.30
	Bas#	$\times 10^9/L$	0.06 ± 0.06	0.27 ± 0.27	0.51 ± 0.51
	RBC	$\times 10^{12}/L$	2.02 ± 0.18	4.09 ± 0.24	4.80 ± 0.50
	HGB	g/L	57 ± 4	130 ± 6	166 ± 8
	HCT	%	17.3 ± 2.0	36.6 ± 3.0	45.2 ± 4.0
	MCV	fL	85.4 ± 5.0	89.5 ± 5.0	94.2 ± 6.0
	MCH	pg	28.2 ± 2.5	31.8 ± 2.5	34.6 ± 2.5
	MCHC	g/L	330 ± 30	355 ± 30	367 ± 30
	RDW-CV	%	14.3 ± 3.0	13.3 ± 3.0	12.9 ± 3.0
	RDW-SD	fL	54.9 ± 10.0	53.5 ± 10.0	55.0 ± 12.0
	PLT	$\times 10^9/L$	59 ± 20	239 ± 40	438 ± 60
MPV	fL	8.9 ± 3.0	8.5 ± 3.0	8.5 ± 3.0	
PDW	fL	10.0 ± 3.0	11.2 ± 3.0	11.4 ± 3.0	
PCT	%	0.053 ± 0.050	0.203 ± 0.100	0.372 ± 0.200	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0126L



BC0126N



BC0126H