



DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-12-26

 2026-03-25

Applicable Instruments	Parameter	Unit	LOT RET0126L	LOT RET0126N	LOT RET0126H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0385 ± 0.0193	0.2446 ± 0.1316	0.2771 ± 0.1200
	RET%	%	0.69 ± 0.35	5.34 ± 2.14	12.57 ± 4.50
	LFR	%	89.7 ± 27.0	89.6 ± 27.0	89.7 ± 27.0
	MFR	%	8.2 ± 12.0	8.4 ± 13.0	8.3 ± 12.0
	HFR	%	2.1 ± 3.0	2.0 ± 3.0	2.0 ± 3.0
	IRF	%	10.3 ± 15.0	10.4 ± 16.0	10.3 ± 15.0
	RHE	pg	16.3 ± 7.0	16.1 ± 6.0	16.2 ± 6.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0126L




RET0126N




RET0126H

DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-12-26

 2026-03-25

Applicable Instruments	Parameter	Unit	LOT RET0126L	LOT RET0126N	LOT RET0126H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0189 ± 0.0151	0.1843 ± 0.0922	0.1727 ± 0.0864
	RET%	%	0.45 ± 0.36	4.54 ± 2.27	8.64 ± 4.32

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0126L



RET0126N



RET0126H