


# Table of Serum Amyloid A Protein Calibrator (SAA) Reference Value

## Serum Amyloid A Protein Calibrator

LOT 2025121501

 2026-12-14

Applicable Instruments	DH-2000T series DH-100T、DH-102T DH-103T、DH-105T DH-106T、DH-108T (Technical File version A2.0 or higher)	DH-800 CS series (Technical File Version A1.3 or higher)
Level (multi-point)	Reference Value (mg/L)	Reference Value (mg/L)
Calibrator 1	10.20	10.50
Calibrator 2	19.90	20.00
Calibrator 3	52.00	57.20
Calibrator 4	115.10	108.00
Calibrator 5	178.09	186.98
Calibrator 6	220.00	215.00
Calibrator 7	267.00	260.00