


# Table of Serum Amyloid A Protein Control (SAA) Reference Value

## Serum Amyloid A Protein Control

LOT 2025121501

 2026-12-14

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH-2000T series DH-100T、DH-102T DH-103T、DH-105T DH-106T、DH-108T (Technical File version A2.0 or higher)	Control 1	8.11	±2.43	DH-800 CS series (Technical File Version A1.3 or higher)	Control 1	7.90	±2.37
	Control 2	52.00	±15.60		Control 2	57.20	±17.16