



Fibrinogen Calibrator Reference Value Table

Fibrinogen Calibrator

LOT 2025121901

 20251219 20271218

Applicable Instruments	Unit	Principle	Reference Value
CA1200 CA1203 CA1205 DC1210 DC1250 DC1280	g/L	Optical method	2.80
	g/L	Magnetic bead	2.92