



DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2026-01-11

 2026-04-10

Applicable Instruments	Parameter	Unit	LOT RET0226L	LOT RET0226N	LOT RET0226H
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	RET#	$\times 10^{12}/L$	0.0375 ± 0.0188	0.1298 ± 0.0779	0.2395 ± 0.1198
	RET%	%	1.07 ± 0.54	3.77 ± 1.51	6.96 ± 2.78
	LFR	%	90.0 ± 27.0	89.9 ± 27.0	90.0 ± 27.0
	MFR	%	6.9 ± 10.0	7.2 ± 11.0	7.0 ± 11.0
	HFR	%	3.1 ± 5.0	2.9 ± 4.0	3.0 ± 5.0
	IRF	%	10.0 ± 15.0	10.1 ± 15.0	10.0 ± 15.0
	RHE	pg	27.3 ± 10.0	27.8 ± 10.0	27.7 ± 10.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0226L



RET0226N



RET0226H