

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT NK0226

M 2026-01-10

H 2026-03-10



Applicable Instruments	Parameter	Unit	Reference Value
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11 or higher)	WBC	$\times 10^9/L$	9.98
	RBC	$\times 10^{12}/L$	4.30
	HGB	g/L	125
	MCV	fL	86.5
	PLT	$\times 10^9/L$	243
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.40
	HGB	g/L	129
	MCV	fL	86.6
	PLT	$\times 10^9/L$	233
DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	126
	MCV	fL	87.4
	PLT	$\times 10^9/L$	221
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	9.08
	RBC	$\times 10^{12}/L$	4.18
	HGB	g/L	125
	MCV	fL	84.2
	PLT	$\times 10^9/L$	240

【NOTE】

- The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0226**
 **2026-01-10**
 **2026-03-10**

Applicable Instruments	Parameter	Unit	Reference Value
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.15
	RBC	$\times 10^{12}/L$	4.20
	HGB	g/L	125
	MCV	fL	85.0
	PLT	$\times 10^9/L$	225
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.38
	RBC	$\times 10^{12}/L$	4.16
	HGB	g/L	126
	MCV	fL	88.8
	PLT	$\times 10^9/L$	224
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	9.92
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	127
	MCV	fL	84.0
	PLT	$\times 10^9/L$	239
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.64
	RBC	$\times 10^{12}/L$	4.39
	HGB	g/L	128
	MCV	fL	84.8
	PLT	$\times 10^9/L$	239


【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 NK0226

 2026-01-10

 2026-03-10



Applicable Instruments	Parameter	Unit	Reference Value
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	10.05
	RBC	$\times 10^{12}/L$	4.40
	HGB	g/L	127
	MCV	fL	83.9
	PLT	$\times 10^9/L$	233
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	9.91
	RBC	$\times 10^{12}/L$	4.51
	HGB	g/L	129
	MCV	fL	85.8
	PLT	$\times 10^9/L$	238
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	8.78
	RBC	$\times 10^{12}/L$	4.30
	HGB	g/L	123
	MCV	fL	85.4
	PLT	$\times 10^9/L$	228
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	9.76
	RBC	$\times 10^{12}/L$	4.32
	HGB	g/L	126
	MCV	fL	84.8
	PLT	$\times 10^9/L$	234

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0226**
 **2026-01-10**
 **2026-03-10**

Applicable Instruments	Parameter	Unit	Reference Value
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.46
	RBC	$\times 10^{12}/L$	4.40
	HGB	g/L	126
	MCV	fL	84.4
	PLT	$\times 10^9/L$	230
UN71 Vet, UN73 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.02
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	128
	MCV	fL	85.3
	PLT	$\times 10^9/L$	236
DM73Vet, DM76Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.17
	RBC	$\times 10^{12}/L$	4.08
	HGB	g/L	123
	MCV	fL	82.7
	PLT	$\times 10^9/L$	265
DM60 Vet, DM61 Vet DM62 Vet, DM63 Vet (Technical File Version A1.4 or higher)	WBC	$\times 10^9/L$	7.85
	RBC	$\times 10^{12}/L$	3.88
	HGB	g/L	121
	MCV	fL	85.5
	PLT	$\times 10^9/L$	233

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.