

DM-CAL

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0226
-----	-------

📅	2026-01-10
---	------------

🕒	2026-03-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.13
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	125
	MCV	fL	90.5
	PLT	$\times 10^9/L$	232
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.20
	RBC	$\times 10^{12}/L$	4.24
	HGB	g/L	126
	MCV	fL	83.0
	PLT	$\times 10^9/L$	237
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.10
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	128
	MCV	fL	89.3
	PLT	$\times 10^9/L$	236
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	8.46
	RBC	$\times 10^{12}/L$	4.15
	HGB	g/L	119
	MCV	fL	88.9
	PLT	$\times 10^9/L$	226

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0226
-----	-------

	2026-01-10
---	------------

	2026-03-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.68
	RBC	$\times 10^{12}/L$	4.30
	HGB	g/L	122
	MCV	fL	86.3
	PLT	$\times 10^9/L$	230
DP-H10, DP-H12 DP-H13, DP-H15 DP-H16, DP-H17 (Technical File Version A3.3 or higher)	WBC	$\times 10^9/L$	9.78
	RBC	$\times 10^{12}/L$	4.85
	HGB	g/L	131
	MCV	fL	84.7
	PLT	$\times 10^9/L$	232
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.46
	RBC	$\times 10^{12}/L$	4.15
	HGB	g/L	129
	MCV	fL	88.9
	PLT	$\times 10^9/L$	226
DP-H20series, DP-H21series DP-H22series, DP-H23series DP-H24series, DP-H25series DP-H26series, DP-H27series DP-H28series, DP-H29series (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.80
	RBC	$\times 10^{12}/L$	4.94
	HGB	g/L	139
	MCV	fL	80.7
	PLT	$\times 10^9/L$	241

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.