

Antithrombin III Calibrator Reference Value Table

AT-III Calibrator

CALIBRATOR

[LOT]2026011901

📊20260119

🕒20280718

Applicable Instrument	Unit	Reference Value
CA1200	%	113.3

【NOTE】

1. The calibrator should be stored in the refrigerator at 2℃~8℃. After reconstitution, the calibrator can be stable for 8 hours at room temperature (10℃~30℃), and can be stable for 12 hours at 2℃~8℃.
2. The calibrator cannot be frozen after reconstitution.
3. The calibrator is for IVD only.