

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT NK0326

M 2026-02-10

H 2026-04-10



| Applicable Instruments | Parameter | Unit | Reference Value |
|--|-----------|--------------------|-----------------|
| D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11 or higher) | WBC | $\times 10^9/L$ | 9.68 |
| | RBC | $\times 10^{12}/L$ | 4.33 |
| | HGB | g/L | 131 |
| | MCV | fL | 89.7 |
| | PLT | $\times 10^9/L$ | 221 |
| UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher) | WBC | $\times 10^9/L$ | 9.36 |
| | RBC | $\times 10^{12}/L$ | 4.43 |
| | HGB | g/L | 134 |
| | MCV | fL | 89.7 |
| | PLT | $\times 10^9/L$ | 209 |
| DH73 Vet (Technical File Version B5.5 or higher) | RBC | $\times 10^{12}/L$ | 4.50 |
| | HGB | g/L | 132 |
| | MCV | fL | 92.1 |
| | PLT | $\times 10^9/L$ | 210 |
| DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher) | WBC | $\times 10^9/L$ | 9.06 |
| | RBC | $\times 10^{12}/L$ | 4.45 |
| | HGB | g/L | 131 |
| | MCV | fL | 87.5 |
| | PLT | $\times 10^9/L$ | 224 |

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0326**
 **2026-02-10**
 **2026-04-10**



| Applicable Instruments | Parameter | Unit | Reference Value |
|---|-----------|--------------------|-----------------|
| DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher) | WBC | $\times 10^9/L$ | 9.12 |
| | RBC | $\times 10^{12}/L$ | 4.32 |
| | HGB | g/L | 132 |
| | MCV | fL | 88.1 |
| | PLT | $\times 10^9/L$ | 209 |
| DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher) | WBC | $\times 10^9/L$ | 9.20 |
| | RBC | $\times 10^{12}/L$ | 4.26 |
| | HGB | g/L | 132 |
| | MCV | fL | 92.3 |
| | PLT | $\times 10^9/L$ | 217 |
| DH71, DH76 (Technical File Version A10.5 or higher) | WBC | $\times 10^9/L$ | 9.55 |
| | RBC | $\times 10^{12}/L$ | 4.48 |
| | HGB | g/L | 131 |
| | MCV | fL | 87.7 |
| | PLT | $\times 10^9/L$ | 221 |
| D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5) | WBC | $\times 10^9/L$ | 9.42 |
| | RBC | $\times 10^{12}/L$ | 4.42 |
| | HGB | g/L | 133 |
| | MCV | fL | 88.5 |
| | PLT | $\times 10^9/L$ | 219 |

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0326**
 **2026-02-10**
 **2026-04-10**



| Applicable Instruments | Parameter | Unit | Reference Value |
|--|-----------|--------------------|-----------------|
| D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher) | WBC | $\times 10^9/L$ | 9.69 |
| | RBC | $\times 10^{12}/L$ | 4.37 |
| | HGB | g/L | 134 |
| | MCV | fL | 88.1 |
| | PLT | $\times 10^9/L$ | 215 |
| D6-CRP, DH76CRP (Technical File Version B2.2 or higher) | WBC | $\times 10^9/L$ | 9.60 |
| | RBC | $\times 10^{12}/L$ | 4.52 |
| | HGB | g/L | 134 |
| | MCV | fL | 89.0 |
| | PLT | $\times 10^9/L$ | 222 |
| DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher) | WBC | $\times 10^9/L$ | 8.58 |
| | RBC | $\times 10^{12}/L$ | 4.36 |
| | HGB | g/L | 129 |
| | MCV | fL | 89.1 |
| | PLT | $\times 10^9/L$ | 216 |

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0326**
 **2026-02-10**
 **2026-04-10**

| Applicable Instruments | Parameter | Unit | Reference Value |
|---|-----------|--------------------|-----------------|
| DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher) | WBC | $\times 10^9/L$ | 9.62 |
| | RBC | $\times 10^{12}/L$ | 4.35 |
| | HGB | g/L | 131 |
| | MCV | fL | 88.9 |
| | PLT | $\times 10^9/L$ | 222 |
| DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher) | WBC | $\times 10^9/L$ | 9.29 |
| | RBC | $\times 10^{12}/L$ | 4.36 |
| | HGB | g/L | 132 |
| | MCV | fL | 87.9 |
| | PLT | $\times 10^9/L$ | 216 |
| UN71 Vet, UN73 Vet (Technical File Version A1.0 or higher) | WBC | $\times 10^9/L$ | 8.91 |
| | RBC | $\times 10^{12}/L$ | 4.40 |
| | HGB | g/L | 133 |
| | MCV | fL | 89.0 |
| | PLT | $\times 10^9/L$ | 219 |

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.