

	文件类型	技术文件	文件版本	V12.3.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-05-en				
文件名称： <p style="text-align: center;"><b>CALIBRATOR N0326 2026-04-10</b></p> 适用范围： 公司所有三分类仪器						
编制：石丽岐 日期：2026.03.02		审核：冯会娟 日期：2026.03.02		批准：冯会娟 日期：2026.03.02		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V12.3.0	/	新建		石丽岐	冯会娟	2026.03.02
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

# DM-CAL

## HEMATOLOGY CALIBRATOR

**Reference Values provided by DYMIND**

<b>CALIBRATOR</b>
-------------------

LOT	N0326
-----	-------

📅	2026-02-10
---	------------

🕒	2026-04-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.05
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	130
	MCV	fL	94.3
	PLT	$\times 10^9/L$	217
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.04
	RBC	$\times 10^{12}/L$	4.26
	HGB	g/L	133
	MCV	fL	86.8
	PLT	$\times 10^9/L$	221
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.57
	HGB	g/L	133
	MCV	fL	92.7
	PLT	$\times 10^9/L$	219
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	8.37
	RBC	$\times 10^{12}/L$	4.20
	HGB	g/L	124
	MCV	fL	92.8
	PLT	$\times 10^9/L$	216

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

# DM-CAL

## HEMATOLOGY CALIBRATOR

**Reference Values provided by DYMIND**

<b>CALIBRATOR</b>
-------------------

LOT	N0326
-----	-------

📅	2026-02-10
---	------------

🕒	2026-04-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.64
	RBC	$\times 10^{12}/L$	4.35
	HGB	g/L	126
	MCV	fL	90.3
	PLT	$\times 10^9/L$	215
DP-H10, DP-H12 DP-H13, DP-H15 DP-H16, DP-H17 (Technical File Version A3.3 or higher)	WBC	$\times 10^9/L$	9.72
	RBC	$\times 10^{12}/L$	4.76
	HGB	g/L	138
	MCV	fL	87.6
	PLT	$\times 10^9/L$	218
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.37
	RBC	$\times 10^{12}/L$	4.18
	HGB	g/L	135
	MCV	fL	92.8
	PLT	$\times 10^9/L$	216
DP-H20series, DP-H21series DP-H22series, DP-H23series DP-H24series, DP-H25series DP-H26series, DP-H27series DP-H28series, DP-H29series (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.64
	RBC	$\times 10^{12}/L$	4.84
	HGB	g/L	145
	MCV	fL	86.1
	PLT	$\times 10^9/L$	228

**【NOTE】**

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.