

	文件类型	技术文件	文件版本	V6.3.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;"><b>CONTROL DM-RET RET0326 2026-05-25</b></p> 适用范围： 1116项目、7003项目						
编制：石丽岐 日期：2026.03.04		审核：冯会娟 日期：2026.03.04		批准：冯会娟 日期：2026.03.04		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V6.3.0	/	新建		石丽岐	冯会娟	2026.03.04
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2026-02-25

 2026-05-25

Applicable Instruments	Parameter	Unit	LOT RET0326L	LOT RET0326N	LOT RET0326H
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0422 $\pm 0.0211$	0.2036 $\pm 0.1222$	0.2149 $\pm 0.1075$
	RET%	%	0.74 $\pm 0.37$	4.29 $\pm 1.72$	9.43 $\pm 3.77$
	LFR	%	89.9 $\pm 27.0$	90.0 $\pm 27.0$	89.4 $\pm 27.0$
	MFR	%	8.0 $\pm 12.0$	8.1 $\pm 12.0$	8.6 $\pm 13.0$
	HFR	%	2.1 $\pm 3.0$	1.9 $\pm 3.0$	2.0 $\pm 3.0$
	IRF	%	10.1 $\pm 15.0$	10.0 $\pm 15.0$	10.6 $\pm 16.0$
	RHE	pg	15.4 $\pm 6.0$	16.3 $\pm 7.0$	16.0 $\pm 6.0$

**【NOTE】**

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0326L



RET0326N




RET0326H


# DM-RET

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**

 2026-02-25

 2026-05-25

Applicable Instruments	Parameter	Unit	LOT RET0326L	LOT RET0326N	LOT RET0326H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0255 $\pm 0.0204$	0.1464 $\pm 0.0732$	0.1337 $\pm 0.0669$
	RET%	%	0.49 $\pm 0.39$	3.33 $\pm 1.67$	6.37 $\pm 3.19$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0326L



RET0326N



RET0326H