



Antithrombin III Control Reference Value Table

AT-III Control

CONTROL

LOT 2026020601

 20260206

 20270505

Applicable Instrument	Unit	Reference Value	Deviation Limit
CA1200 CA1203 CA1205 DC1210 DC1250 DC1280	%	117.9	± 23.6
CA510 CA520 CA530	%	115.7	± 23.1

【NOTE】

1. The control should be stored in the refrigerator at 2°C ~8°C. After reconstitution, the control can be stable for 8 hours at room temperature (10°C~30°C), and can be stable for 12 hours at 2°C~8°C.
2. The control cannot be frozen after reconstitution.
3. The control is for IVD only.